Standard Bidding Document

MS/NDMB&G/2026/42 Procurement of Electro-medical Lab equipment for Bio-genetics Lab-NIASR (NUMS) (Goods)

National

Single Stage-Two Envelope

November 17, 2026

NUMS Procurement Directorate (National University Of Medical Sciences (NUMS)),Office Assistant
ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province).
Phone: +92-320-225-0473, Email: ishaq.wirk@numspak.edu.pk

Table of Contents

REQUEST FOR BIDS 1

Instructions to Bidders 3

Bid Data Sheet 24

Bids Data Sheet (BDS) 25

Eligibility Criteria 30

Evaluation Criteria 31

Items/Lots 33

Related Services of Goods: 40

Items/Lot Specification 41

Price Schedule 55

General Conditions of Contract 57

Special Conditions of Contract 68

Bid Securing Declaration 74

Contract Form 76

Integrity Pact 79

Performance Guarantee Form 81

Annexure 84

Procurement Forms 86

REQUEST FOR BIDS

PROCUREMENT OF GOODS

1. The **NUMS Procurement Directorate (National University Of Medical Sciences (NUMS))** has reserved Funds for the procurement planned for FY **2026-27**. The **NUMS Procurement Directorate (National University Of Medical Sciences (NUMS))** intends to apply part of the proceeds of this Fund to cover eligible payments under the contract for the " **MS/NDMB&G/2026/42 Procurement of Electro-medical Lab equipment for Bio-genetics Lab-NIASR (NUMS)**" with the reference of **"P200268"**

2. The **NUMS Procurement Directorate (National University Of Medical Sciences (NUMS))** invites sealed Bids from eligible Bidders for procurement of goods described in the bidding documents on **EPADS v2.0**.

3. **Single Stage-Two Envelope** will be used by adopting **Quality and Cost Based Selection (QCBS)** Technique for the subject procurement, in line with the Public Procurement Rules, 2026 and any Regulations, Regulatory Guides, Procurement Guidelines or Instructions issued by the Authority from time to time.

4. All Bids must be accompanied by a Bid Security amount described in Bid Security Section in Bidding Document in the form of **Call at Deposit, Demand Draft** or all bids must be accompanied by bid securing declaration in the format specified in the Bidding documents

5. E-Bidding documents, containing detailed terms & conditions, specifications and requirements etc. are available on **e-Pak Acquisition and Disposal System (EPADS)** at **https://epads.gov.pk/opportunities/federal/procurements/200268** for all the interested bidders registered on **EPADS v2.0**. Bidders are required to get themselves registered on **EPADS v2.0** to participate in Bidding process.

6. The e-bids, prepared in accordance with the instructions in the e-Bidding Documents, must be submitted through **EPADS v2.0** on or before **Monday, November 30, 2026 08:30 AM**. E-bids will be opened using **EPADS v2.0** on the same day at **Monday, November 30, 2026 09:00 AM**. Manual

submission of Bids shall not be entertained. Those vendors who have not yet registered on the new version of **EPADS v2.0,** may register themselves on **https://vendors.epads.gov.pk/**. A tutorial to explain the registration process is available at **https://www.youtube.com/watch?v=MNW6T38v7tc**

In terms of Rule 65 of Public Procurement Rules, 2026 Grievance Redressal Committee (GRC) is notified for the subject procurement and notification copy is available on the procuring agency's website and on Authority's website at (**www.ppra.gov.pk**).

NUMS Procurement Directorate (National University Of Medical Sciences (NUMS)), Office Assistant
ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province).
+92-320-225-0473
ishaq.wirk@numspak.edu.pk

Instructions to Bidders

A. Introduction

1.Scope of Bids

1.1 The Procuring Agency (PA), as indicated in the **Bids Data Sheet (BDS)** invites Bids **through EPADS v2.0** for the provision of Goods as specified in the BDS and **in Section V – Evaluation Criteria, Specifications & Schedule of Requirements**. The name, identification, and number of items/deliverables are provided in the **BDS**. The successful Bidders will be expected to provide the goods within the specified period and timeline(s) as stated in the **BDS**.

2. Source of Funds

2.1 Source of funds is referred in Clause-2 I of Invitation for Bids.

3. Eligible Bidders

3.1 A Bidder may be natural person, company or firm or public or semi-public agency of Pakistan or any foreign country**,** or any combination of them with a formal existing agreement (on Judicial Papers) in the form of a joint venture, consortium, or association. In the case of a joint venture, consortium, or association, all members shall be jointly and severally liable for the execution of the Contract in accordance with the terms and conditions of the Contract. The joint venture, consortium, or association shall nominate a Lead Member as nominated in the BDS, who shall have the authority to conduct all business for and on behalf of any and all the members of the joint venture, consortium, or association during the Bidding process, and in case of award of contract, during the execution of the contract.

Verifiable copy of the agreement that forms a joint venture, consortium or association shall be required to be submitted as part of the Bid.

The appointment of Lead Member in the joint venture, consortium, or association shall be confirmed by submission of a valid Power of Attorney to the Procuring Agency.

Any bid submitted by the joint venture, consortium or association shall indicate the part of proposed contract to be performed by each party and each party shall be evaluated (or post qualified if required) with respect to its contribution only, and the responsibilities of each party shall not be substantially altered without prior written approval of the Procuring Agency and in line with any

instructions issued by the Authority.

(The limit on the number of members of JV or Consortium or Association may be prescribed in BDS, in accordance with the guidelines issued by the PPRA).

3.2 The invitation for Bids is open to all prospective suppliers, manufacturers, or authorized agents / dealers subject to any provisions of incorporation or licensing by the respective national incorporating agency or statutory body established for that particular trade or business. Procuring agencies shall specify the registration/licensing requirements for the foreign bidders keeping in view the requirement of that business.

3.3 A Bidder shall not have a conflict of interest. All Bidders found to have a conflict of interest shall be disqualified. A Bidder may be considered to have a conflict of interest with one or more parties in this Bidding process, if they:

1. are associated or have been associated in the past, directly or indirectly with a firm or any of its affiliates which have been engaged by the Procuring Agency to provide consulting services for the preparation of the design, specifications and other documents to be used for the procurement of the Goods to be purchased under this Invitation for Bids.

2. have controlling shareholders in common; or

3. receive or have received any direct or indirect subsidy from any of them; or

4. have the same legal representative for purposes of this Bid; or

5. have a relationship with each other, directly or through common third parties, that puts them in a position to have access to information about or influence on the Bids of another Bidder, or influence the decisions of the Procuring Agency regarding this Bidding process; or

6. Submit more than one Bid in this Bidding process.

3.4 A Bidder may be ineligible if –

1. he is declared bankrupt or, in the case of company or firm, insolvent;

2. payments in favor of the Bidder is suspended in accordance with the judgment of a court of law other than a judgment declaring bankruptcy and resulting (in accordance with the national laws) in the total or partial loss of the right to administer and dispose of its property;

3. the Bidder is convicted, by a final judgment, of any offence involving professional conduct;

4. the Bidder is blacklisted locally or by international organizations and hence debarred due to involvement in corrupt and fraudulent practices, or submission of false or misleading information or performance failure or involved in disrupting procurement process or due to breach of bid Securing Declaration.

3.5 As and when required, bidders shall provide to the Procuring Agency evidence of their eligibility, proof of compliance with the necessary legal requirements to carry out the contract effectively.

3.6 Bidders shall submit Bids relating to the nature, conditions and modalities of sub-contracting wherever the sub-contracting of any elements of the contract amounting to more than ten (10) percent of the Bid price is envisaged.

4. Eligible Goods and Related Services

4.1 All goods and related services to be supplied under the contract shall have their origin in eligible source countries, and all expenditures made under the contract will be limited to such goods and services. For purpose of this Bid, ineligible countries are the countries declared ineligible by the Federal Government.

5. One Bid per Bidder

5.1 A bidder shall submit only one Bid, in the same bidding process, either individually as a Bidder or as a member in a joint venture or any similar arrangement.

5.2 **The Bidder shall not engage a subcontractor for any portion of the contract if the value of such subcontracting exceeds thirty percent (30%) of the total contract amount.**

6. Cost of Bidding

6.1 Any cost incurred by the bidder relating to the preparation and submission of its Bid shall be borne by the bidder, and the Procuring Agency shall in no case be responsible or liable for those costs, regardless of the conduct or outcome of the bidding process.

B. Bidding Documents

7. Contents of Bidding Document

7.1 The Goods required, Bidding procedures, and terms and conditions of the contract are prescribed in the Bidding Documents. In addition to the Invitation for Bids, the Bidding documents which should be read in conjunction with any addenda issued in accordance with **ITB 10.1** include:

Section I -Invitation to Bids

Section II Instructions to Bidders (ITB)

Section III Bid Data Sheet (BDS)

Section IV Evaluation Criteria, Specifications, Schedule of Requirements

Section V Bid Forms

Section VI General Conditions of Contract (GCC)

Section VII Special Conditions of Contract (SCC)

Section VIII Contract Forms

Section IX Integrity Pact

Section X Declaration of Beneficial Ownership Information.

7.2 The Bidder is expected to examine all instructions, forms, terms and specifications in the Bidding documents. Failure to furnish all the information required in the Bidding documents through EPADS v2.0 will be at the Bidder's risk and may result in the rejection of its Bids.

8. Clarification of Bidding documents

8.1 A prospective Bidder requiring any clarification of the Bidding documents may notify the Procuring Agency through EPADS v2.0**.**

8.2 The Procuring Agency will within two (02) working days after receiving the request for clarification, respond to any request for clarification through EPADS v2.0 provided that such request is received not later than three (03) days, prior to the deadline for the submission of Bids as prescribed in **ITB 22**

8.3 The Procuring Agency's response will be forwarded to all identified Prospective Bidders through EPADS v2.0, including a description of the inquiry,

but without identifying its source.

8.4 Should the Procuring Agency deem it necessary to amend the Bidding document as a result of a clarification, it shall do so following the procedure under **ITB 9.**

8.5 If indicated **in the BDS**, the Bidder's designated representative is invited at the Bidder's cost to attend a pre-Bid meeting at the place, date and time mentioned **in the BDS**. During this pre-Bid meeting, prospective Bidders may request clarification of the schedule of requirement, the Evaluation Criteria or any other aspects of the Bidding document.

8.6 Minutes of the pre-Bid meeting, if applicable, including the text of the questions asked by Bidders, including those during the meeting (without identifying the source) and the responses given, together with any responses prepared after the meeting will be uploaded on EPADS v2.0. Any modification to the Bidding documents that may become necessary as a result of the pre-Bid meeting shall be made by the Procuring Agency exclusively through the use of an Addendum pursuant to ITB 9. Non-attendance at the pre-Bid meeting will not be a cause for disqualification of a Bidder.

9. Amendment of Bidding documents

9.1 Before the deadline for submission of Bids, the Procuring Agency for any reason, whether at its own initiative or in response to a clarification requested by a prospective Bidder or Pre-Bid meeting may modify the Bidding documents by issuing addenda through EPADS v2.0.

9.2 The Procuring Agency shall promptly publish the addendum through EPADS v2.0.

Any addendum issued including the notice of any extension of the deadline shall also be communicated through EPADS v2.0 to all the bidders who have already submitted their bids. Such bidders shall have the right to withdraw their already submitted bid and re-submit the revised bid prior to the original or extended bid submission deadline.

9.3 To give prospective Bidders reasonable time in which to take an addendum/corrigendum into account in preparing their Bids, the Procuring Agency may, at its discretion, extend the deadline for the submission of Bids through EPADS v2.0:

Provided that the Procuring Agency shall extend the deadline for submission of Bids, if such an addendum is issued within last three (03) days of the Bids submission deadline.

C. Preparation of Bids

10. Language of Bid

10.1 The Bid prepared by the bidder, as well as all correspondence and documents relating to the Bids exchanged by the Bidder and the Procuring Agency shall be written in the English language unless otherwise specified in the BDS**.** Supporting documents and printed literature furnished by the Bidder may be in another language provided they are accompanied by an accurate translation of the relevant pages in the English language unless otherwise specified in the **BDS**, in which case, for purposes of interpretation of the Bidder, the translation shall govern.

11. Documents and samples Constituting the Bid

11.1 The Bid prepared by the Bidder shall constitute the documents required in the **BDS**.

Details of sample(s) where applicable and requested in the BDS.

Documentary evidence established in accordance with ITB 13 that the Bidder is eligible and/or qualified for the subject bidding process;

Documentary evidence established in accordance with **ITB 03** that the Bidder has been authorized by the manufacturer to deliver the goods into Pakistan, where required and where the supplier is not the manufacturer of those goods;

Documentary evidence established in accordance with ITB 12 that the goods and related services to be supplied by the Bidder are eligible goods and services, and conform to the Bidding Documents;

Bid security or Bid Securing Declaration furnished in accordance with **ITB ---.**

12. Documents Establishing Eligibility of the Goods and Conformity to Bidding documents

12.1 To establish the conformity of the bidder to the Bidding document, the Bidder shall furnish as part of its Bids the documentary evidence that Goods

provided conform to the technical specifications and standards.

13. Documents Establishing Eligibility and Qualification of the Bidder

13.1 Pursuant to **ITB 10,** the Bidder shall furnish, as part of its Bid, all those documents establishing the Bidder's eligibility to participate in the Bidding process and/or its qualification to perform the contract if its Bid is accepted.

14. Form of Bid

14.1 The Bidder shall fill the Form of Bid furnished in the Bidding document. The Bid Form must be completed without any alterations to its format and no substitute shall be accepted.

15. Bid Prices

15.1 The Bid Prices quoted by the Bidder in the Form of Bid and in the Price Schedules shall conform to the requirements specified below or exclusively mentioned hereafter in the Bidding documents.

15.2 **All items in the Schedule of Requirement must be listed and priced separately in the Price Schedule(s). If a Price Schedule shows items listed but not priced and neither explicitly denied, their prices shall be construed to be included in the prices of other items.**

15.3 In the event that any item(s) in the Price Schedule are not quoted, and such omission is determined by the Procuring Agency not to materially affect the responsiveness of the Bid, the Procuring Agency may, for the sole purpose of bid evaluation and comparison, assign a price to the unquoted item(s) in accordance with the following methodology:

(a) The assigned price shall be the average of the prices quoted for the same item(s) by all substantially responsive bidders; or

(b) In cases where no comparable quotations are available, the Procuring Agency shall apply the corresponding item price from its duly approved cost estimate for evaluation purposes.

The prices so assigned shall be utilized exclusively for bid evaluation and comparison, and shall not, under any circumstances, be incorporated into or form part of the Contract Price.

15.4 The Bid price to be quoted in the Form of Bid in accordance with **ITB 13.1** shall be the total price of the Bid.

15.5 The Bidder shall indicate on the appropriate Price Schedule, the unit prices (where applicable) and total Bid price of the Goods it proposes to provide under the contract.

15.6 Prices quoted by the Bidder shall be fixed during the Bidder's performance of the contract and not subject to variation on any account. A Bid submitted with an adjustable price will be treated as non-responsive and shall be rejected.

16. Bids Currencies

16.1 Prices shall be quoted in Pakistani Rupees unless otherwise specified in the BDS in accordance with Rule 42 (2) of the Public Procurement Rules, 2026.

17. Bids Validity Period

17.1 Such extensions may be more than one time but the total period of such extension shall not exceed the original bid validity period.

17.2 Bids shall remain valid for the period specified in the **BDS** after the Bid submission deadline prescribed by the Procuring Agency. A Bid valid for a shorter period shall be rejected by the Procuring Agency as non-responsive**.** The period of Bid validity will be determined from the complementary Bid securing instrument, i.e. the expiry period of Bid Security or Bids Securing Declaration as the case may be.

The procuring agency shall ordinarily be under an obligation to process and evaluate the bid and to issue letter of award within the stipulated bid validity period.

17.3 Under exceptional circumstances, prior to the expiration of the initial Bid validity period, the Procuring Agency may request the Bidders' consent to an extension of the period of validity of their Bids through EPADS v2.0,. The Bid Security provided under **ITB 17** shall also be suitably extended. A Bidder may refuse the request without forfeiting its Bid security or causing to be executed its Bid Securing Declaration. A Bidder agreeing to the request will not be required nor permitted to modify its Bid, but will be required to extend the validity of its Bid Security or Bid Securing Declaration for the period of the extension.

18. Bid Security or Bid Securing Declaration

18.1 The Bidder shall furnish as part of its Bid, a Bid Security in accordance with Rule 36 of the Public Procurement Rules, 2026.

18.2 The original Bid Security shall be enclosed within the sealed envelope and to be submitted physically before closing time for submission of bids. Whereas, scanned copy of bid security shall be uploaded electronically through EPADS v2.0 before closing hours for submission of bids.

18.3 The Bidder who failed to submit the original Bids security before the submission deadline shall be disqualified straightaway.

18.4 The Bid Security or Bid Securing Declaration is required to protect the Procuring Agency against the risk of Bidder’s conduct which would warrant the security’s forfeiture, pursuant to **ITB 18.7.**

18.5 The Bid Security shall be denominated in the local currency or in another freely convertible currency, and it shall be in the form specified in the BDS which shall be in any of the following:

a) a bank guarantee, a bank draft, an irrevocable letter of credit issued by a Scheduled bank in the form provided in the Bidding Documents or another form acceptable to the Procuring Agency;

b) a cashier’s or certified cheque; or

c) another security if indicated in the BDS.

The Bid Security shall be valid for twenty-eight (28) days beyond the end of the validity of the Bid. This shall also apply if the period for Bid Validity is extended. In either case, the form must include the complete name of the Bidder.

18.6 The Bid Security shall be payable promptly upon written demand by the Procuring Agency in case any of the conditions listed in **ITB 18.7** are invoked.

18.7 Unsuccessful Bidders' Bid Security will be discharged or returned as promptly as possible, however in no case later than thirty (30) days after the expiration of the period of Bid Validity prescribed by the procuring agency pursuant to **ITB 17.** The Procuring Agency shall make no claim to the amount of the Bid Security, and shall promptly return the Bid Security documents, after whichever of the following that occurs earliest:

a. the expiry of the Bid Security;

b. the entry into force of a procurement contract and the provision of a Performance Guarantee, for the performance of the contract if such a guarantee, is required by the Bid documents;

c. the rejection by the Procuring Agency of all Bids;

d. cancellation of procurement proceedings.

e. the withdrawal of the Bid prior to the deadline for the submission of Bids, unless the Bid document stipulates that no such withdrawal is permitted.

18.8 The successful Bidder's Bid Security will be discharged upon the Bidder signing the contract**,** or furnishing the Performance Guarantee.

18.9 The Bid Security may be forfeited or the Bid Securing Declaration executed:

a. if a Bidder:

b. withdraws its Bid during the period of Bid Validity as specified by the Procuring Agency, and referred by the Bidder on the Form of Bids except as provided for in **ITB 16.2**; or

c. does not accept the correction of errors**;** or

d. in the case of a successful Bidder, if the Bidder fails:

e. to sign the contract; or

f. to furnish Performance Guarantee**.**

19. Withdrawal, Substitution, and Modification of Bid

19.1 Before Bid submission deadline, any Bidder may withdraw, substitute, or modify its Bid after it has been submitted through EPADS v2.0.

20. Format and Signing of Bid

20.1 The Bidder shall prepare and submit Bids with due diligence after carefully reading all the terms and condition before bid submission deadline through EPADS v2.0.

D. Submission of Bids

21. Submission of Bids through EPADS v2.0 before bid submission deadline.

21.1 The Technical and Financial Bids, shall be submitted electronically through EPADS, before bid submission deadline. The bid submission option shall be automatically disabled once the deadline is over.

22. Deadline for Submission of Bids

22.1 Bids shall be received by the Procuring Agency through EPADS v2.0 before bid submission deadline.

22.2 The Procuring Agency may, under exceptional circumstances, extend the deadline for the submission of Bids, after recording reasons in writing and in an equal opportunity manner.

In such case, all rights and obligations of the Procuring Agency and the Bidders that were previously governed by the original deadline shall thereafter be subject to the revised deadline.

E. Opening and Evaluation of Bids

23. Opening of Bids

23.1 The respective Bid Evaluation Committee of the Procuring Agency shall open all Bids through the EPADS v2.0, on the date and time specified in the Bid Data Sheet (BDS).

23.2 The respective Bid Evaluation Committee **shall generate minutes through EPADS v2.0 containing brief details of bid opening process.** The record of the Bid opening shall include, as a minimum: the name of the Bidder, the Bid price if applicable, and the presence or absence of a Bid Security or Bid Securing Declaration.

23.3 The procuring agency shall live broadcast the opening of bids on national media or on their website or digital channels, if the volume of procurement exceeds five hundred million rupees in case of goods and services and one thousand million rupees in case of works.

23.4 In case the date of opening of bid has been declared as public holiday or the procuring agency fails to open bid due to any EPADS v2.0 related issues, the submission and opening of bids shall be shifted to the next working day at the same time.

23.5 In case of Single Stage One Envelope Procedure, the Bidders names, the Bid prices, the total amount of each Bid and, the presence or absence of Bid Security, Bid Securing Declaration and such other details as the Procuring Agency may consider appropriate, will be announced by the Bid respective Evaluation Committee.

23.6 No bid will be rejected at the time of bid openning except for bids whose original bid security has not been provided to the procuring agency before bid submission deadline.

24. Clarification of Bids

24.1 To assist in the examination, evaluation and comparison of Bids of the Bidders, the respective Bid Evaluation Committee may, ask any Bidder for a clarification of its Bid including breakdown of prices.

24.2 The request for clarification and the response shall be sought through EPADS v2.0 before **three days prior to the deadline for submission of bids.** No change in the prices or substance of the Bids shall be sought, offered, or permitted.

24.3 The alteration or modification in the bids which in any way affect the following parameters will be considered as a change in the substance of a Bid:

1. evaluation & qualification criteria;
2. required scope of work or specifications;
3. all securities requirements;
4. tax requirements;
5. terms and conditions of Bidding documents.
6. change in the ranking of the Bidder

24.4 From the time of bid(s) opening to the time of Contract award if any Bidder wishes to contact the Procuring Agency on any matter related to the Bids it should do so through **EPADS v2.0**.

25. Preliminary Examination of Bids

25.1 Prior to the detailed evaluation of Bids, the respective Bid Evaluation Committee will determine whether each Bid:

1. meets the eligibility criteria defined in **ITB 3**;

2. has been prepared as per the format and contents defined by the Procuring Agency in the Bidding documents;

3. is accompanied by the required securities; and

4. is substantially responsive to the requirements of the Bidding documents.

25.2 The respective Bid Evaluation Committee's determination of a Bid's responsiveness will be based on the contents of the Bid itself.

25.3 A substantially responsive Bid is one which conforms to all the terms, conditions, and specifications of the Bidding documents, without material deviation or reservation. A material deviation or reservation is one that: -

1. affects in any substantial way the scope, quality, or performance of the Goods;

2. limits in any substantial way, inconsistent with the Bidding documents, the Procuring Agency's rights or the Bidders obligations under the Contract; or

3. if rectified, would affect unfairly the competitive position of other Bidders presenting substantially responsive Bids.

25.4 If a Bid is not substantially responsive, it will be rejected by the respective Bid Evaluation Committee and may not subsequently be evaluated for complete technical responsiveness.

26. Examination of Terms and Conditions; Technical Evaluation

26.1 The respective Bid Evaluation Committee shall examine the Bids to confirm that all terms and conditions specified in the **GCC** and the **SCC** have been accepted by the Bidder without any material deviation or reservation.

26.2 The respective Bid Evaluation Committee shall evaluate the technical aspects of the Bids submitted, to confirm that all requirements specified in Schedule of Requirements and Technical Specifications of the Bidding documents have been met without material deviation or reservation.

26.3 If after the examination of the terms and conditions and the technical evaluation, the respective Bid Evaluation Committee determines that the Bid is not substantially responsive in accordance with **ITB,** it shall reject the Bid.

27. Correction of Errors

27.1 Bids determined to be substantially responsive will be checked for any arithmetic errors. Errors will be corrected as follows: -

1. if there is a discrepancy between unit price and the total price that is obtained by multiplying the unit price and quantity, the unit price shall prevail, and the total price shall be corrected, unless in the opinion of the respective Evaluation Committee there is an obvious misplacement of the decimal point in the unit price, in which the total price as quoted shall govern and the unit price shall be corrected;

2. if there is an error in a total corresponding to the addition or subtraction of sub-totals, the sub-totals shall prevail and the total shall be corrected; and

3. where there is a discrepancy between the amounts in figures and in words, the amount in words will govern.

4. Where there is discrepancy between grand total of price schedule and amount mentioned on the Form of Bid, the amount referred in Price Schedule shall be treated as correct subject to elimination of other errors.

27.2 The amount stated in the Bid will be adjusted by the respective Bid Evaluation Committee in accordance with the above procedure for the correction of errors and, with the concurrence of the Bidder, shall be considered as binding upon the Bidder. If the Bidder does not accept the corrected amount, its Bid will then be rejected, and the Bid Security may be forfeited or the Bids Securing Declaration may be executed.

28. Conversion to Single Currency

28.1 For the purposes of comparison of bids quoted in different currencies, the price shall be converted into a single currency in accordance with Rule 42(2) of Public Procurement Rules,2026.

29. Evaluation of Bids

29.1 The Bids, quotations, or proposals shall be evaluated by the respective evaluation committees as per evaluation criteria prescribed in the Bidding Documents in accordance with Rule 11, 12 and 13 of the Public Procurement Rules, 2026.

29.2 While evaluating the bids, the procuring agency shall consider the following:

i. Bids will be evaluated on lump sum bid basis or items wise or Lot wise (contracts), as specified in the BDS; and the Bid Price;

ii. correction of arithmetic errors in accordance with ITB 27;

30. Domestic Preference

30.1 The respective Bid Evaluation Committee shall evaluate and compare bids, allow for preference to domestic bidders, while competing with the international bidders in accordance with the policies of the Federal Government.

The percentage of preference, to be accorded shall be clearly mentioned in the bidding documents under the bid evaluation criteria.

31. Determination of Successful Bid

31.1 **Selection technique will be adopted for determining the Successful Bid in accordance with the criteria referred in the BDS or prescribed in the separate section titled as Evaluation Criteria.**

31.2 **Least Cost Based Selection (LCBS)**
After meeting the requirements of eligibility, qualification and substantial responsiveness, the bid in compliance with all the mandatory (technical) specifications/requirements and/or requisite quality threshold (if any), and having lowest evaluated cost (or financial proposal) shall be considered Successful Bid.

31.3 **Quality and Cost Based Selection (QCBS)**
In such combination, there shall be some specific weightage of both the technical features and financial aspects of the proposal. The financial marks shall be awarded on the basis of inverse proportion calculations. The successful bid shall be declared, on the basis of combined evaluation.

31.4 **Quality Based Selection (QBS)**
After meeting the requirements of eligibility, qualification and substantial responsiveness the bid in compliance with all the mandatory (technical) specifications/requirements and attaining highest marks in the Technical Evaluation considering all other qualitative and/or quantitative parameters (or point rated criteria) for technical proposal(s) such as working methodology, implementation plan, resource allocation, additional functionalities, risk

management approach, knowledge transfer techniques, post implementation methodology etc. shall be treated as highest ranked bid. Later on, the financial proposal of highest ranked bidder shall be opened, however, in case of failure to proceed further with such a bidder, the procuring agency may resort to second highest bidder and so on.

31.5 The Procuring Agency may adopt the Quality & Cost Based Selection Technique due to the following two reasons:

1. Where the Procuring Agency knows about the main features, usage and output of the products; however not clear about the complete features, technical specifications and functionalities of the goods to be procured and requires the bidders to submit their proposals defining those features, specifications and functionalities; or

2. Where the Procuring Agency, in addition to the mandatory requirements and mandatory technical specifications, requires parameters specified in Evaluation Criteria to be evaluated while determining the quality of the goods.

31.6 In such cases, the Procuring Agency may allocate certain weightage to these factors as a part of Evaluation Criteria, and may determine the ranking of the bidders on the basis of combined evaluation in accordance with provisions of Rule 2(1)(xxviii) of the Public Procurement Rules, 2026.

32. Abnormally Low Financial Bids

32.1 An abnormally low bid is a bid that, in light of the Procuring Agency's cost estimate, market conditions, or other received bids, appears to be unrealistically low and may not be sustainable for successfull contract execution.

32.2 **Where the Bid price is considered to be abnormally low, the Procuring Agency shall perform price analysis either during determination of Successful Bids or as a part of the post-qualification process.**

32.3 The Procuring Agency may reject an Abnormally low financial bid in accordance with Rule 43 of the Public Procurement Rules, 2026.

32.4 In order to identify the Abnormally Low Bids (ALB) following approaches can be considered to minimize the scope of subjectivity:

1. Comparing the Bid price with the cost estimate;

2. Comparing the Bid price with the Bid(s) offered by other Bidders submitting substantially responsive Bids; and

3. Comparing the Bids price with prices paid in similar contracts in the recent past either government- or development partner-funded.

32.5 The respective Bid Evaluation Committee will determine to its satisfaction whether the Bidder that is selected as having submitted the successful bid is qualified to perform the contract satisfactorily.

32.6 The determination will take into account the Bidder's financial, technical, and production capabilities. It will be based upon an examination of the documentary evidence of the Bidder's qualifications submitted by the Bidder, as well as such other information as the Procuring Agency deems necessary and appropriate. Factors not included in these Bidding documents shall not be used in the evaluation of the Bidders' qualifications.

32.7 Respective Bid Evaluation Committee may seek "Certificate for Independent Price Determination" from the Bidder and the results of reference checks may be used in determining an award of contract.

Explanation: The Certificate shall be furnished by the Bidder. The Bidder shall certify that the price is determined keeping in view of all the essential aspects such as raw material, its processing, value addition, optimization of resources due to economy of scale, transportation, insurance and margin of profit etc.

32.8 An affirmative determination will be a prerequisite for award of the contract to the Bidder. A negative determination will result in rejection of the Bidder's Bid, in which event the respective Bid Evaluation Committee will proceed to the next ranked Bidder to make a similar determination of that Bidder's capabilities to perform satisfactorily.

F. Award of Contract

33. Criteria of Award

33.1 The Procuring Agency will award the Contract to the Bidder whose Bids has been determined to be substantially responsive to the Bidding documents and who has been declared as Successful Bidder.

34. Negotiations

34.1 In accordance with Rule 52A of Public Procurement Rules, 2026, without changing the cost and scope of work or services, the procuring agency may

negotiate with the successful bidder (with a view to streamline the work or task execution, at the time of contract finalization) on methodology, work plan, staffing and special conditions of the contract.

There shall be no limitation on the successful bidder offering voluntary reduction in bid price without affecting the quality or scope of the procurement.

35. Procuring Agency's Right to reject all bids

35.1 The Procuring Agency reserves the right to reject all bids or proposals at any time prior to the issuance of the Letter of Award, without incurring any liability, in accordance with Rule 45 of the Public Procurement Rules, 2026.

36. Cancellation of Procurement

36.1 The procuring agency reserves the right to cancel the procurement proceedings in accordance with Rule 46 of the Public Procurement Rules, 2026.

37. Procuring Agency’s Right to Vary Quantities at the Time of Award

37.1 The Procuring Agency reserves the right at the time of contract award to increase or decrease the **quantity of** Goods originally specified in these Bidding documents provided this does not exceed **by** 15%, without any change in unit price or other terms and conditions of the Bids and Bidding documents.

38. Notification of Award

38.1 Prior to the award of contract, the procuring agency shall announce and publish the result of bid evaluation on EPADS v2.0 in accordance with Rule 48 of the Public Procurement Rules, 2026.

The Bidder whose Bids has been accepted will be notified of the award by the Procuring Agency prior to expiration of the Bids/Bid Validity period. The Letter of Award will state the sum that the Procuring Agency will pay the successful Bidder in consideration for the delivery of Goods as prescribed by the Contract (hereinafter and in the Contract called the "Contract Price).

38.2 The Letter of award will constitute the formation of the Contract, subject to the Bidder furnishing the Performance Guarantee and signing of the contract.

39. Signing of Contract

39.1 Promptly after issuance of Letter of award, Procuring Agency shall send the successful Bidder the draft Contract, incorporating all terms and conditions as agreed by the parties to the contract.

39.2 Immediately after the Redressal of grievance by the GRC (if any), and **after fulfillment of all conditions precedent** of the Contract Form, the successful Bidder and the Procuring Agency shall sign the Contract.

40. Corrupt & Fraudulent Practices

40.1 Procuring Agencies (including beneficiaries of Government funded projects and procurement) as well as Bidders/Contractors under Government financed contracts, observe the highest standard of ethics during the procurement and execution of such contracts, and will avoid engaging in any corrupt and fraudulent practices.

F. Grievance Redressal & Complaint Review Mechanism

41. Constitution of Grievance Redressal Committee

41.1 The Grievance Redressal Committee shall address the grievance, if any submitted by any party, including the bidder, in accordance with Rule 65 of the Public Procurement Rules, 2026 read with Redressal of Grievances Regulations, 2021.

In case if any party or the bidder is not satisfied with the decision of the GRC or if it fails to decide within ten days, the bidder or the party may file an appeal before the Appellate Committee of the Authority in accordance with Rule 65(7) of the Public Procurement Rules, 2026 to be read with Redressal of Grievances Regulations, 2021.

G. Mechanism of Blacklisting

42. Mechanism of Blacklisting

42.1 The Procuring Agency shall initiate blacklisting proceedings against any bidder, supplier, or contractor in accordance with the Rule 25 of the Public Procurement Rules, 2026.

42.2 The blacklisted/debarred bidder may file the review petition before the Authority in accordance with Rule 25 (6) of the Public Procurement Rules, 2026 to be read with Procedure of filing and disposal of Review Petitions Regulations, 2021.

43. Contract Management

43.1 For each procurement, the Procuring Agency shall designate a Contract Manager, who shall be duly responsible for the efficient management and administration of the contract in strict compliance with the provisions set forth under Rules 58, 59, and 60 of the Public Procurement Rules, 2026.

44. Inspection of goods and evaluation of services.

44.1 The Inspection Committee or the Firm shall have the right to conduct inspections of the goods in accordance with Rule 61 of the Public Procurement Rules, 2026.

45. Socio-economic development.

45.1 In accordance with Rule 63 of the Public Procurement Rules, 2026, preference may be granted to small and medium enterprises and marginalized groups in accordance with the policies notified by the Federal Government.

46. Environmental objectives

46.1 As per Rule 64 of the Public Procurement Rules, 2026, the procuring agency may seek to procure goods with a reduced environmental impact throughout their life cycle when compared to goods services with the same primary function that may otherwise be procured.

Bid Data Sheet

Bids Data Sheet (BDS)

The following specific data for the procurement of Goods to be procured shall complement, supplement, or amend the provisions in the Instructions to Bidders (ITB). Whenever there is a conflict, the provisions herein shall prevail over those in ITB.

BDS Clause Number
ITB Number
Amendments of, and Supplements to, Clauses in the Instruction to Bidders

A. Introduction

BDS Clause Number 1
ITB Number 1.1

Name of Procuring Agency: **NUMS Procurement Directorate (National University Of Medical Sciences (NUMS))**

The subject of procurement is: **MS/NDMB&G/2026/42 Procurement of Electro-medical Lab equipment for Bio-genetics Lab-NIASR (NUMS)**

Period for Provision of Goods:

Extension Period (if applicable):

Expected commencement date: **Thursday, December 17, 2026**

BDS Clause Number 2
ITB Number 2.1

Source of funds [Enter source of funds]

Financial year for the operations of the Procuring Agency: **2026-27**

Name and identification number of the Contract: **P200268**

Identification Number:

BDS Clause Number 3
ITB Clause Number 3.1

Eligible Bidder [Enter eligibility of bidder]

JV/Consortium or Association Allowed: **No**
Number of JV/Consortium Members: **Nil**
see section of eligibility criteria.

B. Bidding Documents

BDS Clause Number 4

ITB Clause Number 4.1

Eligible goods and related services [Enter requirements for eligible goods and related services, if any]

ITB Number 8.1

The Bidders may seek clarifications through **EPADS v2.0** : Clarification Date: Friday, November 27, 2026

ITB Number 8.2

The Bidders may seek clarifications through **EPADS v2.0** : Clarification Date: Friday, November 27, 2026

ITB Number 8.3

The Bidders may seek clarifications through **EPADS v2.0** : Clarification Date: Friday, November 27, 2026

ITB Number 8.5

Pre-bid meeting will be/will not be held [Tick the appropriate] If yes write down the time, and date of the pre-Bid meeting:
[The meeting should take place not later than one weeks before the deadline for Bid submission.]

C. Preparation of Bids

BDS Clause Number 5
ITB Number 10.1

The Language of all correspondences and documents related to the Bids shall be in: **English**

List of documents required along with the bid: No

BDS Clause Number 6
ITB Number 11.1

Documents required / not required with bid submission [enter list of documents] Samples required / not required with the bid [Enter the names and number of samples]

Items/Lots and their related documents:

See section items and Lots

BDS Clause Number 7
ITB Number 12.1

Items / Lots Specifications:

see section of items specifications.

BDS Clause Number 8
ITB Number 15.6

The price shall be **Fixed**.

BDS Clause Number 9
ITB Number 16.1

Currency of the Bids shall be : **PKR**

BDS Clause Number 10
ITB Number 17.1

The Bids/Bid Validity period shall be: **180 Days**

BDS Clause Number 11
ITB Number 18.1

The amount of Bid Security shall be as defined in Bid Security Section for items and lots given in **BDS 6**

Instrument of Bid Security shall be in the Name of : **NUMS Procurement Directorate**

The Bid Security shall be in the form of: **Call at Deposit, Demand Draft**

ITB Number 18.2

The address for submission of original bid security is [Insert the address for bid security submission]

D. Submission of Bids

BDS Clause Number 12
ITB Number 20.1

Bid shall be submitted online on EPADS v2.0 whereas hard copy of the bid security should be submitted to the following;

ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). before bid submission deadline.

Bids that are not submitted on EPADS v2.0 shall be disqualified.

The deadline for Bids submission is: **Monday, November 30, 2026 08:30 AM**

Day *:* ***Monday***

Date: ***Monday, November 30, 2026***

Time *:* ***09:00 AM***

E. Opening and Evaluation of Bids

BDS Clause Number 13
ITB Number 23.1

The Bids opening shall take place on **EPADS v2.0**.

Day *:* ***Monday***

Date: ***Monday, November 30, 2026***

Time *:* ***09:00 AM***

ITB Number 29.2

The procuring agency shall consider the following:
lump sum bid basis; or
items wise; or
Lot wise (contracts),

ITB Number 30.1

BDS Clause Number 14
ITB Number 31.1

Domestic Preference [applicable / not applicable]

ITB Number 31.2

Selection technique adopted will be:
The Procuring Agency shall use any of the following evaluation techniques for determination of successful bid.

Least Cost Base Selection (LCBS)
Quality and Cost Based Selection (QCBS)
Quality Based Selection (QBS)
[Insert the Evaluation Technique]

Selection technique adopted will be: **Quality and Cost Based Selection (QCBS)**
see Evaluation Criteria

F. Review of Procurement Decisions

BDS Clause Number 15
ITB Number 41.1

Grievence against this procurement shall be submitted online on EPADS v2.0.

Arbitrator shall be appointed by mutual consent of the both parties.

ITB Number 42.2

The Address of PPRA Grievance Redressal Appellate Committee to submit a copy of grievance:

Grievance Redressal Appellate Committee,
Public Procurement Regulatory Authority
1st Floor, G-5/2, Islamabad, Pakistan
Tel: +92-51-9202254

Eligibility Criteria

All Active Taxpayers shall be eligible to participate in the procurement process, unless expressly excluded or otherwise restricted under the Eligibility Criteria specified by the Procuring Agency.

Required Registration
NADRA CITIZENSHIP (CNIC/NICOP)
FBR (NTN)
FBR (GSTN)

Eligibility Criteria	Document
All Bidders are advised to must offer quote (Technical & Financial in following format: Item Name Make Model Country of Origin Country of Manufacture Warranty (in years) Expected Delivery Period (in days) after issuance of LoA Quote validity 180 days	Yes

Evaluation Criteria

Eligibile bidder(s) with substantially responsive bid(s) offering **Quality and Cost Based Selection (QCBS)** shall be consider for the award of contract(s).

Quality and Cost Based Selection (QCBS)

Weightage

Technical Evaluation %	Financial Evaluation %
70	30

Technical Marks	100
Passing Marks	**80**
Technical Evaluation Criteria	
Status of Bidder / Profile: (Quantitative)(Doc Required) Manufacturer or Authorized distributor of the Manufacturer **(15)** Public Undertaking, Public Ltd., Private Ltd. Company or Proprietary Firm **(5)**	15
Technical quote / BOQ must be according to following format for each item: Item Name Make Model Country of Origin Country of Manufacture Warranty (in years) Expected Delivery Period (in days) after issuance of LoA Quote validity 180 days (Quantitative)(Doc Required)	20
Technical literature, catalogues, datasheets of offered products original uploaded (Quantitative)(Doc Required)	5
After Sales Service Commitment on letter head for 03 years (Quantitative)(Doc Required)	10

Previous experience for the supply / services with NUMS and other Public sector / Institutions (Quantitative)(Doc Required)	10
Affidavit to the effect that the bidder is not blacklisted by any Govt. agency (Quantitative)(Doc Required)	5
Acceptance of all terms / conditions towards after sales / services as mentioned in the bidding document. (Quantitative) (Doc Required)	5
Financial Evaluation	
One pager bank statement of turnover per year for last three successive years duly certified by concerned bank manager (Qualitative)	20
Quote validity for 180 days on letter head and may be extended if required (Qualitative)(Doc Required)	10

PUBLIC PROCUREMENT REGULATORY AUTHORITY (PPRA) EPADS

Items/Lots

Items Without Lots :

Item	UNSPSC	Delivery Schedule	Quantity	Bid Security	Warranty
Flow Cytometer	Flow cytometer	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	640000 PKR	360 Days
Biomass Extractor: 3- liter Capacity	Extracting equipment for laboratories	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	344720 PKR	720 Days
Fluorescence Microscope	Fluorescent microscopes	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	394760 PKR	360 Days

Item	UNSPSC	Delivery Schedule	Quantity	Bid Security	Warranty
Pharma Crystalizer	Crystallizers	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	383640 PKR	360 Days
Biofermenter	Microbiology fermentation equipment	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	211280 PKR	360 Days
Ultracentrifuge	Ultracentrifuges	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	547113 PKR	360 Days

Item	UNSPSC	Delivery Schedule	Quantity	Bid Security	Warranty
Lyophilizer	Freeze dryers or lyophilizers	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	111200 PKR	360 Days
Gel Dryer	Gel dryers	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	72280 PKR	360 Days
Inverted microscope	Inverted microscopes	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 2/Qty	2/Qty	58380 PKR	360 Days

Item	UNSPSC	Delivery Schedule	Quantity	Bid Security	Warranty
Laminar flow hood	Laminar flow cabinets or stations	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 3/Qty	3/Qty	55600 PKR	360 Days
Short Path Distillation Unit	Laboratory fractional distillation apparatus	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	77840 PKR	360 Days
Vacuum concentrator	Vacuum or centrifugal concentrators	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	100080 PKR	360 Days

Item	UNSPSC	Delivery Schedule	Quantity	Bid Security	Warranty
CO2 Incubators	Water jacketed single chamber three gas incubators	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 4/Qty	4/Qty	61160 PKR	360 Days
Fume hood	Fume hoods or cupboards	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 4/Qty	4/Qty	54210 PKR	360 Days
LIMS/inventory system	Inventory management software	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	33000 PKR	360 Days

Item	UNSPSC	Delivery Schedule	Quantity	Bid Security	Warranty
Biobank Equipment for sample processing	Ultra cold or ultralow upright cabinets or freezers	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	150000 PKR	360 Days
Mobile cryo tank for LN2	Cryogenic or liquid nitrogen freezers	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 7/Qty	7/Qty	14400 PKR	360 Days
Cryogenic Systems	Cryogenic storage boxes	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 2/Qty	2/Qty	328040 PKR	360 Days

Item	UNSPSC	Delivery Schedule	Quantity	Bid Security	Warranty
Autoclave	Steam autoclaves or sterilizers	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	50040 PKR	360 Days
Inverted microscope (with microinjector)	Inverted microscopes	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	355840 PKR	360 Days
Laminar flow hood	Laminar flow cabinets or stations	**Address:** ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province). **Schedule:** For Local item: 02 weeks & Import item; 8-12 weeks **Quantity:** 1/Qty	1/Qty	44480 PKR	360 Days

Related Services of Goods:

No

Items/Lot Specification

Items Without Lots :

Item: Flow Cytometer

UNSPSC: Flow cytometer

Specifications / Requirementsss:

Flow Cytometer: Benchtop flow cytometer Cytometer Machine Cell Counter FCM Flow Cytometer Flow cytometry (FCM) is a technique used to detect and measure physical and chemical characteristics of a population of cells or particles. In this process, a sample containing cells or particles is suspended in a fluid and injected into the flow cytometer instrument. The sample is focused to ideally flow one cell at a time through a laser beam, where the light scattered is characteristic to the cells and their components. Cells are often labeled with fluorescent markers so light is absorbed and then emitted in a band of wavelengths. Tens of thousands of cells can be quickly examined and the data gathered are processed by a computer. Data file/Control Through external computer External computer specification: G5 Desktop System, Intel Core i7, 32 GB RAM, NVMe Solid-State Drive 512 GB SSD Hard Drive or better, CD, DVD writer, 56.6 band Internal Modem; BaseT Ethernet, Audio In, Built-in Ethernet & Transreceiver , dual screen 27” Colour LCD Monitors, Must be provided with latest version of all relevant data analysis softwares with key, Clinical Application: Laser: 488nm solid, Detector: FSC, SSC, FLI-FL5, Fluorescence:FLI: 525/50 (FITC) FL2: 585/40 (PE) FL3: 630/30 (PE-Texas Red/ECD/PI) FL4: 697/60 (PerCP/PE-Cy5/PE-Cy5.5) FL5: 786/60 (PE-Cy7) FTC < 100 MESF; PE < 50 MESF Senstivity a) Analysis and absolute count of lymphocyte subsets: T/B/NK; b) HLA-B27 detection: ankylosing spondylitis; c) Blood disease/lymphoma immunophenotyping; d) Functional lymphocyte subsets analysis: Th1, Treg; e) Count of CD34 hemopoietic stem cell; f) PNH auxiliary diagnosis:CD55 , CD59 , FLARE,CD14 , CD15; g) Multiple myeloma; h) Platelets: count and immunophenotype analysis of platelets; i) Count of Reticulocytes; j) Minor residual disease monitor of acute leukemia; k) DNA ploidy analysis.

Item: Biomass Extractor: 3- liter Capacity

UNSPSC: Extracting equipment for laboratories

Specifications / Requirementsss:

Laboratory-scale biomass extraction system, 1 kg processing capacity, 3-litre production column, quick-switch column configuration, rapid solvent recovery system, high-wattage silicone heating system with four heaters and three independent temperature controllers, integrated suction/drying tank with transparent and stainless-steel lids, solvent-compatible construction, controlled heating and solvent recovery, assembled and factory-tested unit, supplied with extraction column, heating assemblies, controllers, suction/drying tank, connecting accessories, operating manual, installation and operational training, minimum one-year warranty, country of origin North America, Europe or UK, subject to manufacturer documentation.

Item: Fluorescence Microscope

UNSPSC: Fluorescent microscopes

Specifications / Requirementsss:

Microscope Stages and Stage equipment: • Three point mounted regular fixed stage 248 mm x 212 mm. Aluminum, ceramic-coated, Scratch-proof, precisely plane-parallel, including a round 88 mm insert with 10 mm opening. • Object guide for regular, heatable/ cooling fixed stages, with a size of 248 mm x >200 mm, , deep-lying coaxial control drive. Travel Range: 83 x 127 mm. • Universal Holding Frame M, suitable for fixed stages with object guide. For Petri dishes with diameter 24 mm up to 68 mm and slides with a length up to 120 mm. Objectives: • HC PL Fluotar 4x/0.13 PH0 • HC PL FLUOTAR 10x/0.32 PH1 • HC PL FLUOTAR 20x/0.55 PH2 • HC PL FLUOTAR 40x/0.80 PH2 • HC PL FLUOTAR 100x/1.32 OIL PH3 • HC PL FL L 63x/0.70 CORR PH2 Camera: • 6.3 MP camera Includes o color CMOS sensor o Pixel size 2.4 µm x 2.4 µmColor filter RGB (Bayer) o Color depth 3x12 bits o Exposure time 1 ms - 1 s o Dynamic range >72 dB o Fast live image 3072 x 2048 pixel with up to 15 frames per sec o Gain 1x - 251x o Supported operating systems Win11/Win 10 o Interface: USB3 o DVD with instruction for use o Start guide. • C-mount adapter 0.55x HC, for 1/2" and 0.55" cameras Computer System • Operating System: Win11 Pro64 • Processor: Intel Core i7-13700 2.1GHz, 16Cores • Main Memory: 32GB (2x16GB) DDR5 4800 ECC RAM • Graphics Board: NVIDIA RTX A2000 12GB 4mDP GFX • System Drive: 512GB PCIe 2280 M.2 SSD • Temp Drive: 1TB PCIe 2280 M.2 SSD • Data Drive: 4TB 7200 SATA Enterprise 3.5” HDD • Ethernet: 1Gbit • 32 Inch LED Screen • Wireless Mouse and Keyboard UPS 2KVA

Item: Pharma Crystalizer

UNSPSC: Crystallizers

Specifications / Requirementsss:

Pharmaceutical Crystallizer: Laboratory-scale pharmaceutical crystallization system for batch or continuous crystallization, Product Features: 1. Integrated PLC control 2. Large particle products can be produced through a designed fine crystal elimination system 3. Automatic material discharge, automatic crystallizer opening/closing, automatic heating/cooling system, online concentration and particle size monitoring and control, integrated PLC control. It can be applied to QbC (Quality by Control) experiments and production processes. Product Advantages: 1. Uniform stirring, controlling primary nucleation: Prevents rapid burst nucleation, avoids unnecessary crystal formation, reduces crystal agglomeration, product wall adhesion, and optimizes crystal particle size. 2. Reduces secondary nucleation: Low stirring rate reduces the intensity of collisions between the stirrer and solid, solid and vessel wall, and between solids, thus reducing secondary nucleation. 3. Uniform product particle size, improving product appearance (crystal habit), and reducing batch-to-batch differences in crystal habit during intermittent production. 4. Reduces differences in crystallization process and product caused by scale-up during pilot-scale to large-scale production. 5. Reduces scaling during production; improves particle agglomeration problems. Product Specifications: 1 m^3-90 m^3 Product Materials: Stainless steel, PTFE-lined, titanium, etc.

Item: Biofermenter

UNSPSC: Microbiology fermentation equipment

Specifications / Requirementsss:

Biofermenter: Control Tower: Weight 13.5 kg | 29.7 lbs: Footprint bench surface (W + H + D) 200 + 495 + 300 mm | 7.9 + 19.5 + 11.8 in: Required space incl. connectors, pump heads etc. (W + H + D): 320 + 495 + 475 mm | 12.6 + 19.5 + 18.7 in: Power supply 100 to 240 VAC 50/60Hz, 8 A: Housing Metal housing, coated: Operating Temperature 0 °C – 40 °C: Maximum relative humidity 80% for temperatures up to 31 °C decreasing linearly to 50% relative humidity at 40 °C: Motor drive – Maintenance-free: – Quiet direct drive: – 150 W: Motor rpm: direct coupling: 1 L glass: 30 – 1,400 rpm: 2 L glass: 30 – 1,100 rpm: 5 L glass: 30 – 800 rpm: 2 L SU: 30 – 400 rpm: Interfaces – 2 +

external signal inputs 4 – 20 mA: – Ethernet: – Potential-free alarm contact: – USB: – Interface to chiller: – Interface to external pump: Regulatory compliance CE, NRTL recognized by the U.S. OSHA (Occupational Safety and Health Administration): System requirements for operation by tablet and smartphone: iOS 7 or higher; Android 4 or higher Integrated Pumps Quantity 3: Controller Fixed speed (regulated on/off) : Speed 43 rpm: Pump head Watson Marlow 114, Fast Load pump head: Flow rates: (tube wall thickness 1.6 mm) ID 0.5 mm: 0.8 ml/min: ID 0.8 mm: 1.7 ml/min: ID 1.6 mm: 6.0 ml/min: ID 2.4 mm: 12.5 ml/min: ID 3.2 mm: 20.2 ml/min: External Pumps: Quantity 1 (only in the Advanced package): Controller Regulated rpms: Speed Max. 200 rpm: Pump head Watson Marlow 120, Fast Load pump head: Pump Module: Base Unit: Probes and Controllers UniVessel® Glass UniVessel® SU: Temperature, separate sensor Temperature control 0 – 40°C: Display resolution 0.1°C • Temperature integrated in pH probe Temperature control 0 – 60°C ;Display resolution 0.1°C • DO probe, reusable Polarographic probe: Digital communication with Control Tower: Range: 0 –100% : Display resolution: 0.1% : • DO sensor, single-use DO sensor patch: Range: 0 –100% air saturation: Display resolution: 0.1% air saturation: • pH probe, reusable Combination electrode, digital communication with Control Tower: Range: 2 – 12 pH: Display resolution: 0.01 pH: • Process Control and Measurements: Gas inlet – Gas pressure 1.5 barg (21.76 psig), ±10%: – Gases: dry, oil and dust-free: – Gas inlet connections with hose barb for 6 + 3 mm (0.24 x 0.12 in.) reinforced tubing: Gas outlet – Gas outlet pressure: max. 0.8 barg (11.6 psig) – Outlets to the culture vessel with hose barb for 3.2 + 1.6 mm (0.12 x 0.06 in.) silicone tubing: Two-gas System for Microbial Fermentations: Continuous and automatic aeration control for air and O2: Gas flow control units Total of two; one per gas segment: Flow rates of the gas flow control units (all vessel sizes) : Air and O2: 100 – 7,500 ccm: Measuring and control accuracy of the gas flow control units ± 5% full scale: Four-gas System for Cell Culture Application: Continuous and automatic aeration control for air, O2, N2 and CO2: Gas flow control units Total of four; one per gas segment: Flow rates of the gas flow control units (all vessel sizes): Air and N2: 10 – 500 ccm | O2 and CO2: 5 – 250 ccm: Measuring and control accuracy of the gas flow control units ± 5% full scale Aeration Module Cooling (only included in the BIOSTAT® A for microbial fermentation) Chiller with automatic cooling water valves for recirculated cooling of the culture vessel and exhaust cooler via cooling finger Weight 19.5 kg | 43 lbs. Dimensions (W + H + D) 220 + 495 + 325 mm | 8.6 + 19.49 + 12.8 in. Volume of cooling water tank 1.5 L: Power supply – BB-8822002: 230VAC, 50/60 Hz, T2.5 A - BB-8822003: 115VAC, 60 Hz, T4A: Supply temperature 8°C ±2°C at an ambient temperature of 21°C Cooling capacity 150 W at an ambient temperature of 21°C: International protection rating IP21 (protected from solid objects > 0.5 inch and vertically dripping water) Cooling water hardness Max. 12 dH: Connection with cooling finger via quick-connect coupling Heating Temperature control with electric heating jacket for culture vessel Electrical filter heating for exhaust air filter (only included in the BIOSTAT® A for cell culture applications): Culture vessel temperature control range (See above) Heating power – 1 L Glass: 66 W: - 2 L Glass: 112 W: - 5 L Glass: 264 W: - 2 L Single-use: 132 W: Country of origin (COI): Northern America, UK, Europe or equivalent Laboratory Biofermenter (Reference Brand: Sartorius, Germany, BIOSTAT® A): Bench-top bioreactor/fermenter system with interchangeable 1 L, 2 L and 5 L glass vessels and 2 L single-use vessel option, coated metal control tower weighing approximately 13.5 kg, maintenance-free direct-drive motor rated 150 W, adjustable agitation speed of 30–1,400 rpm depending on vessel configuration, three integrated peristaltic pumps and provision for an additional regulated external pump, digital temperature control with 0.1°C resolution, pH measurement range 2–12 with 0.01 pH resolution, dissolved oxygen measurement range 0–100%, automatic aeration control, two-gas microbial fermentation s

Item: Ultracentrifuge

UNSPSC: Ultracentrifuges

Specifications / Requirementsss:

The system should be a refrigerated floor-standing ultracentrifuge with a maximum speed of RCF 100,000 x g or more, supplied with a fixed-angle rotor (8 x 50 mL). The speed range should be 500 to 29,000 rpm with ±25 rpm accuracy, and the temperature range should be -20 to +40 °C with ±2 °C accuracy and a pre-cooling function and greencool technology using a CO_2 natural-refrigerant cooling system (GWP = 1). The rotor must be locked and identified automatically (Auto-Lock and Auto-ID) to prevent over-speeding. The drive shall be a high-torque brushless, imbalance-tolerant system (5% of opposing loads) with a Smart Vacuum system. The unit shall offer 9 acceleration and 10 deceleration profiles with a brake-off option, a run time up to 99 h 59 min 59 s plus HOLD, 30 step-run triplets and an ACE integrator. Control shall be via a glove-friendly touchscreen with 120 programs, 120 password-protected users with lock-out, a multilingual interface, rotor calculator, onboard training videos, automatic door opening, and USB and Ethernet run logs. It should have Biocontainment lids and auto lock rotor exchange in as little as 3 sec, Noise shall be below 57 dBA with fixed-angle rotors. Should have anti-vibration rubber boots and wheel locks. The equipment MUST be supplied with power backup (APC or any reputed international brand UPS of 20kVA or more with dry batteries for uninterrupted power supply during the operation) Relevant Certifications include CE, IEC 61010-1, 61010-2-020 and 61010-2-011, RoHS and WEEE, with ISO 6 cleanroom compatibility. Country of Origin: USA, Europe, Germany or Western Europe. Standard: CE marking or equivalent international grading.

Item: Lyophilizer

UNSPSC: Freeze dryers or lyophilizers

Specifications / Requirementsss:

Extremely compact while incorporating robust technical features and outstanding control performances. The Telstar LyoQuest benchtop freeze dryer has been developed to meet the most demanding requirements of a research freeze dryer. Outstanding features Extremely compact bench mounted table-top unit. Large capacities of condensation (6 kg/24 h) in the minimum possible space. Two different process oriented control systems, both PLC based, adapts the unit to every requirement: - Basic control for LyoQuest

Series: allows automatic pressure control and shelves temperature control. - Plus control for LyoQuest Plus Series: allows automatic pressure control and shelves temperature programming. Touch screen that displays system operation parameters and alarm messages. Descriptive and timely information display and setting selection is provided. Monitoring software (LyoLogger) to connect unit to PC available. Ice condenser capacity: (8 (6 kg/24h), Final condenser temperature*: <-85 C, Numbner of compressors: 2, Total Power: 1.4kW, Standard stoppering chamber with 8 port manifold: Cylindrical acrylic chamber with 3 heated shelves (for vials and serum), manual vial closing device and 8 port manifold with valves (for flasks or wide-neck filter bottles with CN29/32 connection) ISo, CE certififed COuntry of origin: Europe, UK and USA

Item: Gel Dryer

UNSPSC: Gel dryers

Specifications / Requirementsss:

Western Blot Apparatrus Specifications: The western blot system should include tank, lid with power cables, electrode assembly, casting stand for minimum 4 gels, five 10-well combs of 1.0mm thickness, and Should be able to transfer 4 small gels in one run. The system should be capable of blot transfer in less than 10 minutes. Electroblotting core/module. or laboratory vacuum gel drying system for drying polyacrylamide and sequencing gels, Specifications: minimum drying surface of 35 × 45 cm and capacity for 16 mini gels, nine Criterion-size gels, two 20 × 20 cm large-format gels, or one sequencing gel. The system shall feature variable temperature control from 50°C to 90°C, three programmable drying cycles, a porous polypropylene gel support plate, transparent silicone sealing gasket, and vacuum-assisted drying to minimize gel cracking. The system shall include a compatible environmentally friendly, water-based vacuum pump without vacuum pump oil, with vacuum performance greater than 29 in Hg at startup and a flow rate of approximately 21 L/min at 50 Hz. The complete package shall include vacuum tubing, connectors, drain tubing, gel support, sealing gasket, filter-paper backing, cellophane-membrane backing, sequencing-gel filter paper, and operating manuals. Electrical supply: 220–240 V AC, 50/60 Hz. Supplier shall confirm compatibility, all accessories, warranty, installation, training, spare-parts availability and after-sales support.

Item: Inverted microscope

UNSPSC: Inverted microscopes

Specifications / Requirementsss:

Inverted microscope with Camera supporting brightfield, apodized phase contrast, phase contrast, Emboss Contrast and upgradeable epi-fluorescence. Binocular tube with 35° inclination; ergonomic Siedentopf-type tube with 15–35° or wider inclination and interpupillary distance of 50–75 mm. Eyepieces: 10× (FN 22), 12.5× (FN 16) and 15× (FN 14.5). Coarse focusing stroke of 5.0 mm per rotation and fine focusing stroke of 0.1 mm per rotation. Sextuple revolving nosepiece, multi-module contrast condenser, and stage compatible with culture dishes, slides, haemocytometers and multiwell plates. Optional: Epi-fluorescence system compatible with multiple LED excitation wavelengths and filter cubes. Power supply: 220–240 V AC, 50 Hz. Installation, training, one year warranty and after-sales support included.

Item: Laminar flow hood

UNSPSC: Laminar flow cabinets or stations

Specifications / Requirementsss:

Specifications: BSC Cabinets Class-II, Type A2 with metal stand and wheels with lock. Working chamber: 4 feet with Dual HEPA/ULPA filters, including an exhaust fan with thimble and Vertical Laminar Air Flow through a HEPA H14 filter generating a vertical laminar airflow (downflow) that creates an ISO 14644-1 (Class 5/ GMP (Grade A) work area, preventing ambient air entering the work area, ensuring product protection. Microprocessor controlled Digital display, with running hours couter.• The manually operated front sliding safety glass window which should be completely closed to protect the work are whilst the cabinet is not in operation. Cabinets made of stove-enamel coated steel • Polished AISI-304 stainless steel working surface • HEPA filter protected with laminar screen to provide incomparable uniform flow • U.V. resistant tempered glass side walls • Sliding front safety glass window • Control panel with alarms and parameters: Digital control board with LCD display* including: • ON/OFF button and light indicator • Selection of air speed: working or stand-by • Button for U.V. operation with possibility of exposure time programming • LEDs on control panel show status of: lighting, U.V., clogging of filter • Alarm LED • Display with real-time air speed indication • Elapsed time counter for the fan and U.V. light Key Internal dimensions WxDxH (mm):1230x693x615 or more Air velocity m/s: 0.36 Consumption W: 123 Lighting Lux: ≥1500 Noise level (according to ISO 11201) dB (A): 55 Main Filter HEPA H14 Efficiency: Efficiency 99.995%, EN 1822 by MPPS method (99,999% D.O.P. test @ 0.3µm particles) Pre-filter Efficiency EN 779: G3 Fans EC High efficiency Centrifugal. Coating Powder coating based on polyester resins Working Surface AISI-304

Item: Short Path Distillation Unit

UNSPSC: Laboratory fractional distillation apparatus

Specifications / Requirementsss:

Short Path Distillation Unit: UIC KDL5 or equivalent laboratory-scale short-path distillation system. Item 1. Feed System: 1) Jacketed feed tank with flanged lid and spare nozzles. 2) Jacketed gear pump designed for 300°C, with speed adjustable through a frequency inverter. 3) Jacketed feed line to the evaporator. Item 2. Short-Path Evaporator: 1) Maximum evaporator temperature: 300°C. 2) Borosilicate glass construction. 3) Internal condenser coil. 4) Wiper system with three strands or equivalent. 5) Digital display of wiper speed at the motor drive. 6) Operating pressure down to 0.001 mbar, depending on product and operating conditions. 7) Jacketed feed, residue and distillate nozzles. Item 3. Distillate and Residue Discharge Pumps (2 pumps): 1) Jacketed gear pumps designed for 300°C, with speed adjustment through frequency inverters. 2) Jacketed discharge siphons to provide pressure isolation from the atmosphere. Item 4. Vacuum System: 1) Oil-sealed rotary-vane vacuum pump. 2) Oil diffusion pump for achieving high vacuum, including vacuum oil. 3) Flanged cold trap with condensate collection vessel. 4) Dosing valve for pressure adjustment and venting. 5) Digital vacuum-pressure gauge. 6) All required vacuum connection pipes and hoses. Item 5. Heating and Cooling: 1) Independent heating units for the evaporator, feed tank, residue and distillate lines, and internal condenser. 2) Independent cooling unit for the cold trap. Item 6. Rack/Frame: 1) Movable stainless-steel frame with connection clamps and four wheels. Item 7. Power Supply and Control System: 1) Voltage: 220–240 V AC. 2) Frequency: 50/60 Hz, compatible with local supply. 3) Single-phase power distribution panel. 4) Separate control panel for pumps and temperature control. Installation, commissioning, training, warranty and after-sales support included. Country of origin (COI): North America, UK, Europe or equivalent. (replaces Sr. 18 of Annex-IX (C2)

Item: Vacuum concentrator

UNSPSC: Vacuum or centrifugal concentrators

Specifications / Requirementsss:

Vacuum Concentrator: The Micro-CenVac (NB-503CIR) is a compact centrifugal vacuum concentrator from N-Biotek designed for rapid sample evaporation in molecular biology, genomics, proteomics, and drug discovery laboratories. Key Features & Specifications • Infrared (IR) Lid Heating: Features a special infrared-emitting glass lid that directly heats the sample from above to accelerate evaporation rates well beyond standard concentrators. • Real-Scope Viewing: Allows continuous real-time visual inspection of the sample volume while it is spinning at high speed. • Vacuum Pump: Includes a built-in, oil-free, Teflon-coated diaphragm vacuum pump (reaching up to 100 mbar). • Drive System: Maintenance-free magnetic induction drive for quiet, low-vibration operation. • Temperature Control: PID-controlled heating ranging from ambient +5°C to 70°C, alongside a standby pre-heating function. • Capacity Options: Accommodates various rotor configurations, including microcentrifuge tubes (e.g., 24 x 1.5 ml, up to 30ml x 8, 15ml x 12, or 10ml x 16. USA/UK/Korea and Europe

---

Item: CO2 Incubators

UNSPSC: Water jacketed single chamber three gas incubators

Specifications / Requirementsss:

Microprocessor controlled 150 Liter or more, waterjacketed, stackable CO2 Incubator with Tri-gas option (O2, N2 and CO2). MUST be provided with one touch Dry Heat Sterilization protocol. Temperature control from 4°C above ambient to 50°C, with control accuracy ±0.1°C. with high Tem. - It should have six-sided direct heating with fanless, gentle convection circulation to provide stable temperature control, excellent uniformity and rapid recovery with no over shoot. - It should have CO2 control range from 0.2 to 20% with control accuracy and uniformity of ±0.1% and should have rapid recovery of at least 0.7% per minutes. - It should have Infrared (IR) CO2 sensor It should come with minimum 4 adjustable height shelves & humidity reservoir (removable) to achieve at least 95% RH. K21- It should have independent door heater eliminate condensation on inner door surface. - It should have HEPA filter on CO2 inlet; It should have 25mm access port. - It should be ISO 9001 or CE Certification or equivalent. Country of origin (COI): Northern America, UK, Europe or equivalent

---

Item: Fume hood

UNSPSC: Fume hoods or cupboards

Specifications / Requirementsss:

Fume Hood: (Esco Frontier Mono or equivalent) Industrial-grade support frame constructed of electro-galvanized steel with oven-baked epoxy-polyester powder coating. Internal baffle system made of chemical-resistant phenolic resin laminate for uniform airflow. Bypass airflow system to maintain effective airflow when the sash is lowered. Stainless steel Grade 304 airfoil designed to sweep the work surface. Main body made of electro-galvanized steel with corrosion-resistant epoxy-polyester coating. Sash: vertical-rising, tempered laminated and framed safety glass. Complete unit with LED lighting, airflow monitoring, base storage cabinet and suitable exhaust blower system. Electrical supply: 220–240 V AC, 50 Hz. Compliance with applicable laboratory fume hood standards, including EN 14175 or equivalent. Country of origin (COI): North America, UK, Europe or equivalent (replaces Sr. 26 of Annex-IX (C2)).

Item: LIMS/inventory system

UNSPSC: Inventory management software

Specifications / Requirementsss:

LabWare LIMS or Thermo Scientific SampleManager LIMS, or equivalent internationally recognized system. Features: sample registration and tracking; laboratory inventory management (chemicals, reagents and consumables); stock quantity and expiry-date monitoring; batch/lot tracking; storage-location mapping; barcode/QR-code support; low-stock alerts; instrument integration; equipment maintenance and calibration records; configurable workflows; results management; reporting and dashboards; role-based access; audit trails; data backup and recovery. Cloud-based or on-premises deployment. Supply includes software licensing, installation, configuration, training, technical support and maintenance.

Item: Biobank Equipment for sample processing

UNSPSC: Ultra cold or ultralow upright cabinets or freezers

Specifications / Requirementsss:

Ultra-Low Temperature Freezer −80°C (One piece): Upright type, Capacity approximately 505 L, Temperature setting range: −40°C to −90°C, PL ADV Control System Microprocessor-based temperature control with digital display and 0.1°C resolution. microprocessor-based temperature controller with digital display LED display with 5-keys capacitive keyboard, with 0.1°C resolution, adjustable temperature settings, high/low-temperature, door-open and power-failure alarms (audible and visual) with battery backup for controller and alarm, insulated cabinet with polyurethane foam 140 mm on sides / 140 mm rear / 110 mm outer door and magnetic door gasket, lockable door, minimum 4-5 adjustable stainless steel or powder-coated shelves (height-adjustable) for compartmentalization of samples, Static cooling, with 2-stage cascade refrigeration system; air-cooled condenser with filterless design. reagents and kits, MUST be supplied with additional storage baskets, dividers or racks for sub-compartmentalization, individual inner doors or door-mounted compartments to minimize cold loss on opening where available, uniform temperature distribution, automatic temperature monitoring with USB or RS-485 data logging port, CFC-free and HFC-free refrigerant, Continuous temperature monitoring and data logging, high/low temperature alarms, door-open and power-failure alarms, audible and visual alerts, and USB or equivalent data export. Lockable external door. Power supply: 220–240 V AC, 50 Hz.

Item: Mobile cryo tank for LN2

UNSPSC: Cryogenic or liquid nitrogen freezers

Specifications / Requirementsss:

Mobile cryotank with all reuired accessories. LN2 storage capacity: 130 L; storage capacity: 4,800 × 2 mL cryovials; 6 storage racks, 8 shelves per rack, compatible 100-vial cryoboxes; vacuum-insulated, high-strength aluminium construction; lockable insulated lid; static holding time: approximately 153 days; evaporation rate: approximately 0.85 L/day. Dimensions: approximately 892 mm height × 683 mm diameter; empty weight: approximately 41 kg. Supplied with compatible storage racks, cryoboxes, and heavy-duty mobile roller base with castors. Low-level LN2 alarm with remote monitoring, optional. Country of origin: North America, UK, Europe or equivalent.

Item: Cryogenic Systems

UNSPSC: Cryogenic storage boxes

Specifications / Requirementsss:

LN2 Cryobank with capacity of 200 Liters to be supplied with 03x LN2 supply tanks each bearing 240 Liters or more). Microprocessor-controlled with Battery Back-Up Kit, Platform Risers Required for Racks in Vapor Phase Storage 4 Inch, Square Racks for Cardboard boxes, Vapor Phase Rack should include cardboard boxes and cardboard dividers for 1.5/2 mL vials

---

Item: Autoclave

UNSPSC: Steam autoclaves or sterilizers

Specifications / Requirementsss:

Digital steam sterilization autoclave; Type: Vertical or horizontal, electrically heated; Chamber capacity: 50–100 litres; Chamber dimensions: Approximately Ø 35–45 cm × 50–70 cm deep for a vertical model; Chamber material: SS 304 stainless steel; SS 316; Outer body: Corrosion-resistant stainless steel or powder-coated steel; Operating temperature Selectable validated cycles, typically 121°C and 134°C; Operating pressure: Approximately 1.1 bar gauge at 121°C; approximately 2.0 bar gauge at 134°C; Temperature control: Digital microprocessor/PID controller; Pressure monitoring: Digital pressure display or gauge; Timer: Programmable sterilization holding time; Safety features: Door-lock interlock, overpressure protection, over-temperature cut-off, low-water protection; Sterilization cycles: Suitable cycles for solid instruments, compatible materials, and approved waste loads; Drying function: for wrapped instruments required drying; Water supply:Manual filling or automatic water feed, according to model; Validation: Temperature and pressure verification; provision for biological and chemical indicators; Power supply: 220–240 V AC, 50 Hz, subject to heater rating and electrical installation; Accessories: Sterilization baskets, trays, drain arrangement, and operating manual; Warranty: Minimum one year

---

Item: Inverted microscope (with microinjector)

UNSPSC: Inverted microscopes

Specifications / Requirementsss:

Trinocular inverted biological microscope with programable microinjector: Optical system: Infinity-corrected optical system; Viewing head: Trinocular head, 45° inclined, adjustable interpupillary distance; Eyepieces: WF 10×, field number ≥20 mm; Objectives: 4×, 10×, 20× and 40× long-working-distance objectives; Total magnification: 40×–400× with 10× eyepieces; Phase contrast: Phase-contrast capability, preferably with 10×, 20× and 40× objectives; Condenser: Long-working-distance condenser suitable for cell culture vessels; Stage: Mechanical stage, approximately 200 × 240 mm or equivalent; Focusing: Coaxial coarse and fine focusing with tension adjustment; Illumination: LED transmitted-light illumination with adjustable intensity; Camera: Digital camera, preferably 5 MP or higher, with image-capture software; Specimen compatibility: Tissue-culture flasks, multiwell plates, Petri dishes and slides, subject to stage clearance; Power supply: 220–240 V AC, 50 Hz; Accessories: Phase rings, compatible culture-vessel holders, dust cover and operating manual; Warranty: Minimum 1 year; 2 years preferred. • Microscope Must be equiped with programmable microinjector with digital display and should come with injection capillary and capillary holder • Capable of pronuclear and cytoplasmic microinjection and micromanipulation of primates, mouse and zebrafish oocytes. Capable of consistent microinjection of as low as picolitres of liquids. • Come with Integrated pressure supply as well as connection for external pressure supply. • Micromanipulator with 3 axes movements controllable with joystick. • Micromanipulator with electronic connectivity for PC and inverted microscope. • Micromanipulator angle adjustment from 0-90 degree. • • Micromanipulator microinjection systems should be able to work as one unit. Warranty: Minimum 01 year, preferred two years. Additional supply of compatible online UPS (with batteries) for smooth working is highly desirable.

Item: Laminar flow hood

UNSPSC: Laminar flow cabinets or stations

Specifications / Requirementsss:

Digital steam sterilization autoclave; Type: Vertical or horizontal, electrically heated; Chamber capacity: 50–100 litres; Chamber dimensions: Approximately Ø 35–45 cm ×

50–70 cm deep for a vertical model; Chamber material: SS 304 stainless steel; SS 316; Outer body: Corrosion-resistant stainless steel or powder-coated steel; Operating temperature Selectable validated cycles, typically 121°C and 134°C; Operating pressure: Approximately 1.1 bar gauge at 121°C; approximately 2.0 bar gauge at 134°C; Temperature control: Digital microprocessor/PID controller; Pressure monitoring: Digital pressure display or gauge; Timer: Programmable sterilization holding time; Safety features: Door-lock interlock, overpressure protection, over-temperature cut-off, low-water protection; Sterilization cycles: Suitable cycles for solid instruments, compatible materials, and approved waste loads; Drying function: for wrapped instruments required drying; Water supply:Manual filling or automatic water feed, according to model; Validation: Temperature and pressure verification; provision for biological and chemical indicators; Power supply: 220–240 V AC, 50 Hz, subject to heater rating and electrical installation; Accessories: Sterilization baskets, trays, drain arrangement, and operating manual; Warranty: Minimum one year

PUBLIC PROCUREMENT REGULATORY AUTHORITY (PPRA) EPADS

Price Schedule

For Individual Items

#	Item Title	Quantity	Unit Price (PKR)	Total Price (PKR)	Delivery Location	Delivery Period / Year	Country of Origin
1							
2							

For Lots

#	Lot Title	Total Lot Price (PKR)	Country of Origin
1	[Lot 1 Title]		

PUBLIC PROCUREMENT REGULATORY AUTHORITY
(PPRA)
EPADS

General Conditions of Contract

A. General

1. Definitions

1.1 Unless the context otherwise requires, the following terms whenever used in this Contract shall have the same meaning and shall be interpreted as indicated

a. “Applicable Law” means the laws and any other instruments having the force of law in the Government’s Country, or in such other country as may be specified in the Special Conditions of the Contract (SC), as they may be issued and in force from time to time;

b. The “Arbitrator” is the person appointed with mutual consent of both the parties, to resolve contractual disputes as provided for in the General Conditions of the Contract GCC Clause 35 hereunder.

c. “Authority” means Public Procurement Regulatory Authority.

d. The “Commencement Date” is the date when the Supplier shall commence execution of the contract as specified in the SCC.

e. “Completion date” means the date of completion of the Contract by the Bidder as certified by the Procuring Agency.

f. “Procuring Agency” means:-

a. any Ministry, Division, Department or any Office of the Government;

b. any authority, corporation, body or organization established by or under a Law or which is owned or controlled by the Government;.

g. “The Contract” means an agreement enforceable by law;

h. “The Contract Price” means the price payable to the Bidder under the Contract for the full and proper performance of its contractual obligations;

i. “Ancillary Services” means those services ancillary to the provision of Goods, such as transportation and insurance, and any other incidental services, such as installation, commissioning, provision of technical assistance, training, and other such obligations of the Bidder covered under the Contract;

j. “GCC” means the General Conditions of Contract contained in this section;

k. “SCC” means the Special Conditions of Contract by which the GCC may be amended or supplemented;

l. **“**Day” means calendar day unless indicated otherwise.

m. “Defective Goods” are those goods which are below standards, requirements or specifications stated by the Contract.

n. “Delivery” means the transfer of the goods from the supplier equipment, machinery, and /or other materials which the Supplier is required to supply to the Procuring Agency under Contract.

o. “Effective Date” means the date on which this Contract comes into force and effect.

p. “Force Majeure” means an unforeseeable event which is beyond reasonable control of either Party and which makes a Party’s performance of its obligations under the Contract impossible or so impractical as to be considered impossible under the circumstances.

For the purposes of this Contract, “Force Majeure” means an event which is beyond the reasonable control of a Party, is not foreseeable, is unavoidable, and its origin is not due to negligence or lack of care on the part of a Party, and which makes a Party’s performance of its obligations hereunder impossible or so impractical as reasonably to be considered impossible in the circumstances. and includes, but is not limited to, war, riots, civil disorder, earthquake, fire, explosion, storm, flood, epidemics, or other adverse weather conditions, strikes, lockouts or other industrial action (except where such strikes, lockouts or other industrial action are within the power of the Party invoking Force Majeure to prevent), confiscation or any other action by Government agencies.

q. The Bidder” means the individual or corporate body whose Bids to provide the Goods has been accepted by the Procuring Agency;

r. “The Project Site,” where applicable, means the place or places named in Bids Data Sheet and technical Specifications;

s. “Government” means the Government of Pakistan;

t. “Subcontractor” means any entity to which the Bidder subcontracts any part of the Goods.

u. The “Services” means the work to be performed by the Bidder pursuant to this Contract and as prescribed in the Specifications and Schedule of Activities included in the Bidder’s Bids;

v. “Party” means the Procuring Agency or the Bidder, as the case may be, and “Parties” means both of them;

w. “Foreign Currency” means any currency other than the currency of the country of the Procuring Agency;

x. “Completion Date” means the date of completion of the contract by the Bidder as certified by the Procuring Agency;

y. “In Writing” means communicated in written form with proof of receipt;

z. “Local Currency” means the currency of Pakistan;

a. “Origin” means the place where the Goods were mined, grown, or produced or from which the Services are supplied. Goods are produced when, through manufacturing, processing, or substantial and major assembly of components, a commercially recognized new produce results that is substantially different in basic characteristics or in purpose or utility from its components.

b. “Specification” means the Specification of the Goods and performance of incidental services in accordance with the relevant standards included in the Contract and any modification or addition made or approved by the Procuring Agency.

c. “Procuring Agency” means: -
any Ministry, Division, Department or any Office of the Federal Government;
any authority, corporation, body or organization established by or under a Federal Law or which is owned or controlled by the Federal Government;

2. Application and Interpretation

2.1 These General Conditions shall apply to the extent that they are not superseded by provisions of other parts of the Contract.

2.2 In interpreting these Conditions of Contract headings and marginal notes are used for convenience only and shall not affect their interpretations unless specifically stated; references to singular include the plural and vice versa; and masculine include the feminine. Words have their ordinary meaning under the language of the Contract unless specifically defined.

3. Applicable Law

3.1 The contract shall be governed and interpreted in accordance with the laws of Pakistan, unless otherwise specified in SCC.

4. Governing Language

4.1 The Contract as well as all correspondence and documents relating to the Contract exchanged between the Bidder and the Procuring Agency, shall be written in the **English language** unless otherwise stated in the **SCC**. Supporting documents and printed literature that are part of the Contract may be in another language provided these are accompanied by an accurate translation of the relevant passages in English, in which case, for purposes of interpretation of the Contract, this translation shall govern.

5. Notices

5.1 Any notice, request, or consent made pursuant to this Contract shall be in writing and shall be deemed to have been made when delivered in person to an authorized representative of the Party to whom the communication is addressed, or when sent by registered mail, telex, telegram, or facsimile to such Party at the address specified in the **SCC**.

6. Standard

6.1 The Goods supplied under this Contract shall conform to the standards mentioned in the Technical Specifications, and, when no applicable standard is mentioned, the American Standards (such as ACI, IEEE, ASME, etc.) or the Pakistani standards such as PSQCA Such standards shall be the latest issued by the concerned institution.

7. Delivery/Location

7.1 The Goods shall be delivered to such locations as the Procuring Agency may approve and as specified in SCC.

8. Authorized Representatives / Authority of Member in charge

8.1 Any action required or permitted to be taken, and any document required or permitted to be executed, under this Contract by the Procuring Agency or the Bidder may be taken or executed by the officials specified in the **SCC**.

B. Commencement, Completion, Modification, and Termination of Contract

8. Effectiveness of Contract

8.1 This Contract shall come into effect on the date the Contract is signed by both parties and such other later date as may be stated in the SCC.

9. Commencement of the Contract

9.1 The Bidder shall confirm availability of Key Experts and begin carrying out the Services not later than the number of days after the Effective Date specified in the SCC.

10. Program

10.1 Before commencement of the Contract, the Bidder shall submit to the Procuring Agency for approval a Program showing the general methods, arrangements, order and timing for all activities. The Services shall be carried out in accordance with the approved Program as updated.

11. Starting Date/Expiration Date

11.1 The Bidder shall start carrying out the Services Five (05) days after the date the Contract becomes effective, or at such other date as may be specified in the **SCC.**

11.2 Unless terminated earlier pursuant to Clause **GCC 16** hereof, this Contract shall expire at the end of such time period after the Effective Date as specified in the SCC.

12. Entire Agreement

12.1 This Contract contains all covenants, stipulations and provisions agreed by the Parties. No agent or representative of either Party has authority to make, and the Parties shall not be bound by or be liable for, any statement, representation, promise or agreement not set forth herein.

13. Modification

13.1 Any modification or variation of the terms and conditions of this Contract, including any modification or variation of the scope, may only be made by written agreement between the Parties. However, each Party shall give due consideration to any Bids for modification or variation made by the other Party.

13.2 In cases of any modifications or variations, the prior written consent of the Procuring Agency is required.

14. Force Majeure

14.1 Definition

For the purposes of this Contract, "Force Majeure" means an event which is beyond the reasonable control of a Party and which makes a Party's performance of its obligations under the Contract impossible or so impractical as to be considered impossible under the circumstances.

14.2 No Breach of Contract

The failure of a Party to fulfill any of its obligations under the contract shall not be considered to be a breach of, or default under, this Contract in so far as such inability arises from an event of Force Majeure, provided that the Party affected by such an event (a) has taken all reasonable precautions, due care and reasonable alternative measures in order to carry out the terms and conditions of this Contract, and (b) has informed the other Party as soon as possible about the occurrence of such an event.

14.3 Extension of Time

Any period within which a Party shall, pursuant to this Contract, complete any action or task, shall be extended for a period equal to the time during which such Party was unable to perform such action as a result of Force Majeure. However, such period shall not exceed sixty (60) days.

14.4 Payments

During the period of their inability to perform the Services as a result of an event of Force Majeure, the Bidder shall be entitled to continue to be paid under the terms of this Contract, as well as to be reimbursed for additional costs reasonably and necessarily incurred by them during such period for the purposes of the Services and in reactivating the Service after the end of such period.

15. Termination

15.1 By the Procuring Agency

The Procuring Agency may terminate this Contract in case of the occurrence of any of the events specified in paragraphs (a) through (e) of this Clause. In such an occurrence the Procuring Agency shall give at least thirty (30) calendar days' written notice of termination to the Bidder in case of the events referred to in (a) through (d); at least sixty (60) calendar days' written notice in case of the event referred to in (e);

a. If the Bidder fails to remedy a failure in the performance of its obligations hereunder, as specified in a notice of suspension;

b. If the Bidder becomes (or, if the Bidder consists of more than one entity, if any of its members becomes) insolvent or bankrupt or enter into any agreements with their creditors for relief of debt or take advantage of any law for the benefit of debtors or go into liquidation or receivership whether compulsory or voluntary;

c. If the Bidder fails to comply with any final decision reached as a result of arbitration proceedings;

d. If, as the result of Force Majeure, the Bidder is unable to perform a material portion of the Services for a period of not less than sixty (60) calendar days;

e. If the Procuring Agency, in its sole discretion and for any reason whatsoever, decides to terminate this Contract;

15.2 By the Bidder

The Bidder may terminate this Contract, by not less than thirty (30) calendar days' written notice to the Procuring Agency, in case of the occurrence of any of the events specified in paragraphs (a) through (d) of this Clause.

a. If the Procuring Agency fails to pay any money due to the Bidder pursuant to this Contract and not subject to dispute within forty-five (45) calendar days after receiving written notice from the Bidder that such payment is overdue.

b. If, as the result of Force Majeure, the Bidder is unable to perform a material portion of the Services for a period of not less than sixty (60) calendar days.

c. If the Procuring Agency fails to comply with any final decision reached as a result of arbitration.

d. If the Procuring Agency is in material breach of its obligations pursuant to this Contract and has not remedied the same within forty-five (45) days (or such longer period as the Bidder may have subsequently approved in writing) following the receipt by the Procuring Agency of the Bidder's notice specifying such breach.

C. Obligations of the Bidder

16. General

16.1 Standard of Performance

1. The Bidder shall deliver the product and carry out the Services with all due diligence, efficiency and economy, in accordance with generally accepted professional standards and practices, and shall observe sound management practices, and employ appropriate technology and safe and effective equipment, machinery, materials and methods. The Bidder shall always act, in respect of any matter relating to this Contract or to the Services, as a faithful adviser to the Procuring Agency, and shall at all times support and safeguard the Procuring Agency's legitimate interests in any dealings with the third parties.

16.2 Law Applicable to Goods

The Bidder shall deliver the goods in accordance with the Contract and in accordance with the Law of Pakistan and shall take all practicable steps to ensure that any of its Experts and Sub-Bidders, comply with the Applicable Law.

17. Conflict of Interests

17.1 Bidder Not to Benefit from Commissions and Discounts.

The remuneration of the Bidder shall constitute the Bidder's sole remuneration in connection with this Contract or the Services, and the Bidder shall not accept for their own benefit any trade commission, discount, or similar payment in connection with activities pursuant to this Contract or to the Services or in the discharge of their obligations under the Contract, and the Bidder shall use their best efforts to ensure that the Personnel, any Subcontractors, and agents of either of them similarly shall not receive any such additional remuneration.

17.2 Bidder and Affiliates Not to be Otherwise Interested in Project

The Bidder agree that, during the term of this Contract and after its termination, the Bidder and its affiliates, as well as any Subcontractor and any of its affiliates, shall be disqualified from providing Goods for any project resulting from or closely related to the Services.

17.3 Prohibition of Conflicting Activities

Neither the Bidder nor its Subcontractors nor the Personnel shall engage, either directly or indirectly, in any of the following activities:

a. during the term of this Contract, any business or professional activities in the Government's country which would conflict with the activities assigned to them under this Contract;

b. during the term of this Contract, neither the Bidder nor their Subcontractors shall hire public employees in active duty or on any type of leave, to perform any activity under this Contract;

18. Confidentiality

18.1 Except with the prior written consent of the Procuring Agency, the Bidder and the Experts shall not at any time communicate to any person or entity any confidential information acquired in the course of the contract.

19. Insurance to be Taken Out by the Bidder

19.1 The Bidder(a) shall take out and maintain, and shall cause any Subcontractors to take out and maintain, at its (or the Subcontractors', as the case may be) own cost but on terms and conditions approved by the Procuring Agency, insurance against the risks, loss or damage, and for the coverage, as shall be specified in the SCC; and (b) at the Procuring Agency's request, shall provide evidence to the Procuring Agency showing

that such insurance has been taken out and maintained and that the current premiums have been paid.

20. Bidder's Actions Requiring Procuring Agency's Prior Approval

20.1 The Bidder shall obtain the Procuring Agency's prior approval in writing before taking any of the following actions:

(a) appointing such members of the Personnel not provided by the Bidder;

(b) changing the Program of activities; and

(c) any other action that may be specified in the SCC.

21. Reporting Obligations

21.1 The Bidder shall submit to the Procuring Agency the reports and documents in the numbers, and within the periods as prescribed by the Procuring Agency.

22. Liquidated Damages

22.1 If the Supplier fails to deliver any or all of the Goods or to perform the Services within the period(s) specified in the Contract, the Procuring Agency shall, without prejudice to its other remedies under the Contract, deduct from the Contract Price, as liquidated damages, a sum equivalent to the percentage specified in **SCC** of the delivered price of the delayed Goods or unperformed Services for each week or part thereof of delay until actual delivery or performance, up to a maximum deduction of the performance security (or guarantee) specified in **SCC**. Once the said maximum is reached, the Procuring Agency may consider termination of the Contract pursuant to **GCC Clause 16.**

22.2 Correction for Over-payment

If the Intended Completion Date is extended after liquidated damages have been paid, the Procuring Agency shall correct any overpayment of liquidated damages by the Bidder by adjusting the next payment certificate. The Bidder shall be paid interest on the overpayment, calculated from the date of payment to the date of repayment, at the rates specified in SCC.

22.3 Lack of performance penalty

If the Bidder has not corrected a Defect within the time specified in the Procuring Agency's notice, a penalty for Lack of performance will be paid by the Bidder. The amount to be paid will be calculated as a percentage of the cost of having the Defect corrected, assessed as specified in the SCC.

23. Performance Guarantee

23.1 The Performance Security (or Guarantee) shall be provided to the Procuring Agency no later than the date specified in the Letter of Acceptance and shall be issued in an amount and form acceptable to the Procuring Agency as specified in SCC, and denominated in the types and proportions of the currencies in which the Contract Price is payable as specified in the SCC.

23.2 Within Seven (07) days from the issuance of acceptance letter from the Procuring Agency, the successful Bidder shall furnish the Performance Guarantee in shape of ------- at the discretion of the PA in the amount **specified in SCC.** In case the amount of Bids security is equal or greater than

23.3 The proceeds of the Performance Guarantee shall be payable to the Procuring agency as compensation for any loss resulting from the Supplier's failure to complete its obligations under the Contract.

23.4 The Performance Guarantee shall be denominated in the currency of the Contract, or in a freely convertible currency acceptable to the Procuring agency and shall be in the acceptable form as specified in SCC.

23.5 The Performance Guarantee will be discharged by the Procuring agency and returned to the Supplier not later than thirty (30) days following the date of completion of the Supplier's performance obligations under the Contract, including any warranty obligations, unless otherwise **specified in SCC.**

23.6 The bid security of the successful bidder shall not be substituted with the Performance Guarantee.

24. Fraud and Corruption

24.1 The Procuring Agency requires the Supplier to disclose any commissions or fees that may have been paid or are to be paid to agents or any other party with respect to the Bidding process or execution of the Contract. The information disclosed must include at least the name and address of the agent or other party, the amount and currency, and the purpose of the commission, gratuity or fee.

25. Sustainable Procurement

25.1 The Bidder shall conform to the sustainable procurement contractual provisions, if and as specified in the SCC.

D. Bidder's Personnel

26. Description of Personnel

26.1 The titles, agreed job descriptions, minimum qualifications, and estimated periods of engagement in the carrying out of the Services of the Bidder's Key Personnel. The Key Personnel listed by title as well as by name are hereby approved by the Procuring Agency.

27. Removal and/or Replacement of Personnel

27.1 Except as the Procuring Agency may otherwise agree, no changes shall be made in the Key Personnel. If, for any reason beyond the reasonable control of the Bidder, it becomes necessary to replace any of the Key Personnel, the Bidder shall provide as a replacement a person of equivalent or better qualifications.

27.2 If the Procuring Agency finds that any of the Personnel have (i) committed serious misconduct or have been charged with having committed a criminal action, or (ii) have reasonable cause to be dissatisfied with the performance of any of the Personnel, then the Bidder shall, at the Procuring Agency's written request specifying the grounds thereof, provide as a replacement a person with qualifications and experience acceptable to the Procuring Agency.

27.3 The Bidder shall have no claim for additional costs arising out of or incidental to any removal and/or replacement of Personnel.

E. Obligations of the Procuring Agency

28. Assistance and Exemptions

28.1 The Procuring Agency shall use its best efforts to ensure that the Government shall provide the Bidder such assistance and exemptions as specified in the SCC.

29. Change in the Applicable Law

29.1 If, after the date of this Contract, there is any change in the Applicable Law with respect to taxes and duties which increases or decreases the cost of the related Services rendered by the Bidder, then the remuneration and reimbursable expenses otherwise payable to the Bidder under this Contract shall be increased or decreased accordingly by agreement between the Parties, and corresponding adjustments shall be made to the amounts referred in the SCC.

30. Services and Facilities

30.1 The Procuring Agency shall make available to the Bidder and the Experts, for the purposes of the Services and free of any charge, the services, facilities and property described , at the times and in the manner specified in the SCC or terms of reference.

30.2 In case that such services, facilities and property shall not be made available to the Bidder, the Parties shall agree on (i) any time extension that it may be appropriate to grant to the Bidder for the performance of the Services, (ii) the manner in which the Bidder shall procure any such services, facilities and property from other sources, and (iii) the additional payments, if any, to be made to the Bidder as a result thereof.

F. Payments to the Bidder

31. Contract Price

31.1 The price payable shall be in Pakistani Rupees unless otherwise specified in the SCC. Prices charged by the Supplier for Goods delivered under the Contract shall not vary from the prices quoted by the Supplier in its Bid.

32. Terms and Conditions of Payment

32.1 Payments will be made to the Bidder according to the payment schedule stated in the SCC and as per actual invoice submitted by the Bidder.

32.2 Unless otherwise stated in the SCC, the advance payment shall be made against the provision by the Bidder of a bank guarantee for the same amount, and shall be valid for the period stated in the SCC. Any other payment shall be made after the conditions listed in the SCC for such payment have been met, and the Bidder have submitted an invoice to the Procuring Agency specifying the amount due.

33. Currency of Payment

33.1 Any payment under this Contract shall be made in the currency(ies) specified in the SCC.

G. Quality Control

34. Identifying Defects

34.1 The principle and modalities of Inspection of the Goods by the Procuring Agency shall be as indicated in the SCC. The Procuring Agency shall check the Bidder’s performance and notify him of any Defects that are found. Such checking shall not affect the Bidder’s responsibilities. The Procuring Agency may instruct the Bidder to search for a Defect and to uncover and test any service that the Procuring Agency considers may have a Defect. Defect Liability Period is as defined in the SCC.

35. Correction of Defects, and

Lack of Performance Penalty

35.1 The Procuring Agency shall give notice to the Bidder of any Defects before the end of the Contract. The Defects liability period shall be extended for as long as Defects remain to be corrected.

35.2 Every time notice a Defect is given, the Bidder shall correct the notified Defect within the length of time specified by the Procuring Agency's notice.

35.3 If the Bidder has not corrected a Defect within the time specified in the Procuring Agency's notice, the Procuring Agency will assess the cost of having the Defect corrected, the Bidder will pay this amount, and a Penalty for Lack of Performance.

36. Taxes and Duties

36.1 A Supplier shall be entirely responsible for all taxes, duties, fees, etc., incurred until delivery of the contracted Goods to the Procuring Agency.

H. Settlement of Disputes

37. Alternate Dispute Resolution

37.1 Upon the commencement of the Procurement Contract, any disputes arising between the contracting parties shall be resolved in accordance with Rule 66 of the Public Procurement Rules, 2026.

37.2 The procuring agency may refer the matter to the Chief Justice Islamabad High Court or Managing Director PPRA or the Secretary Ministry of Law & Justice for appointment of Arbitrator.

37.3 The fee for the Arbitrator shall be specified in Pak Rupees as determined by the appointing authority which shall be borne and shared equally by the contracting parties.

38. Change Orders

The Procuring Agency may at any time, by a written order given to the Supplier, make changes within the general scope of the Contract in any one or more of the following:

a)Drawings, designs, or specifications, where Goods to be furnished under the Contract are to be specifically manufactured for the Procuring Agency;
b)The method of shipment or packing;
c)The place of delivery; and/or
d)The Services to be provided by the Supplier.

39. Contract Amendenments

39.1 Subject to GCC Clause 32, no variation in or modification of the terms of the Contract shall be made except by written amendment signed by the parties.

40. Sub-Contracts

40.1 The Supplier shall consult the Procuring Agency in the event of subcontracting under this contract if not already specified in the Bid. Subcontracting shall not alter the Supplier's obligations.

Special Conditions of Contract

SECTION VIII. SPECIAL CONDITIONS OF CONTRACT

The following Special Conditions of Contract shall supplement the General Conditions of Contract. Whenever there is a conflict, the provisions herein shall prevail over those in the Conditions of Contract. The corresponding clause number of the GCC is indicated in parentheses.

Number of GC Clause

Amendments of, and Supplements to, Clauses in the General Conditions of Contract

Number of GC Clause 1

Definitions

The Procuring Agency is: **NUMS Procurement Directorate (National University Of Medical Sciences (NUMS)), Office Assistant ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province).**

The Supplier is:

The title of the subject procurement is: **MS/NDMB&G/2026/42 Procurement of Electro-medical Lab equipment for Bio-genetics Lab-NIASR (NUMS)**

Number of GC Clause 3

Applicable/Governing Law:

The Contract shall be interpreted in accordance with the laws of Islamic Republic of Pakistan

Number of GC Clause 4

Language:

The language of the Contract, all correspondence and communications to be given, and all other documentation to be prepared and supplied under the Contract shall be in **English.**

Number of GC Clause 5

Notices:

The addresses for the notices are:

The Authority:

NUMS Procurement Directorate (National University Of Medical Sciences (NUMS)), Office Assistant ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province).
+92-320-225-0473
ishaq.wirk@numspak.edu.pk

Contractor/ Bidder:

[Name, address and telephone number].

The Contractor/ Bidder's Representative(s)

[Name, address, telephone number and e-mail address]

Number of GC Clause 7.1

The Authorized Representatives are:

For the Procuring Agency:

NUMS Procurement Directorate (National University Of Medical Sciences (NUMS)), Office Assistant ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province).
+92-320-225-0473
ishaq.wirk@numspak.edu.pk

For the Bidder:

Name: ……

Designation: ……..

Address: ……..

Number of GC Clause 8

Effectiveness of the contract

Number of GC Clause 9

Commencement of Contract:

Number of GC Clause 11.2

Expiration of Contract:

Number of GC Clause 15

Termination

In the event of termination of the contract due to any reason as already defined in the General Conditions of Contract, the Bidder shall be responsible for providing to the Authority the Goods till the time of alternate arrangements.

Number of GC Clause 17

Conflict of Interest:

The Procuring Agency reserves the right to determine on a case-by-case basis whether the Bidder should be disqualified from providing goods or services due to a conflict of a nature described in Clause GCC 17**.**

Number of GC Clause 22

Liquidated Damages

If the Bidder fails to provide services as required under the contract or in case of any data loss/data breach or any incident compromising the data security or other such failures related to any services, the Bidder shall pay to the Procuring Agency as Liquidated Damages at a rate of **1.00%** to **10.00%** of the Contract value, in accordance with the extent of performance failure & the cost of investigating such incidents as judged by the Authority.

Number of GC Clause 23

Performance Guarantee:

Instrument of Performance Guarantee shall be in the Name of : **NUMS Procurement Directorate**

The amount of performance guarantee shall be **2.00%** of the contract price in acceptable form of **Call at Deposit**

Number of GC Clause 32

Payment terms:

Payment will be made to the Bidder against the procured Goods and services according to the actual invoice or running bills submitted by the Bidder against the services provided within the time given in the conditions of the contract.

Number of GC Clause 33

Currency of Payment:

All the payment to be released to the contractor/Bidder shall be in Pakistani Rupees.

Number of GC Clause 34

Identifying Defects:

The Authority reserves the right at any time to inspect the premises of the provider to inspect the goods and monitor the goods being provided.

Inspections & Tests Requirements

For being Brand New, bearing relevant reference numbers of the equipment (Certificate from supplier)

For Physical Fitness having No Damages (Certificate from supplier)

For the Country of Origin as quoted by the Supplier (Certificate from manufacturer)

For conformance to specifications and performance parameters, through Prior to delivery inspection (Inspection Report by Procurement Committee / Inspection Team)

For successful operation at site after complete installation, testing and commissioning of the equipment (Installation, Testing and Commissioning Report by Procurement Committee / Inspection Team)

1. Authorised distibutor letter2. GD documents at the time of delivery3. IQ, OQ PQ certificates from manufacture verifiable from issuance authority4. ISO -13485 verifiable from issuance authority

Delivery & Documents

Copies of the Supplier's invoice showing Goods' description, quantity, unit price, and total amount;

Original and two copies of the usual transport document (for example, a negotiable bill of lading, a non-negotiable sea waybill, an inland waterway document, an air waybill, a railway consignment note, a road consignment note, or a multimodal transport document) which the buyer may require to take the goods;

Copies of the packing list identifying contents of each package;

Insurance Certificate;

Manufacturer's or Supplier's Valid Warranty Certificate;

Inspection Certificate issued by the Nominated Inspection Agency (if any), and the Supplier's Factory Inspection Report;

Certificate of Origin.

The above documents would be required even if the equipment has already been imported and is available with the supplier ex-stock

Number of GC Clause 37

Following is the guidance for Dispute Resolution

(a) For Contracts to be entered with foreign Contractor/ Service Provider:

All disputes arising in connection with the present Contract shall be finally settled under the Rules of Conciliation and Arbitration of the International Chamber of Commerce by one or more arbitrators appointed in accordance with said Rules.

(b) For Contracts to be entered with nationals of Pakistan:

i. If any dispute of any kind whatsoever shall arise between the Authority and the Bidder in connection with or arising out of the Contract, including without prejudice to the generality of foregoing, any question regarding its existence, validity, termination and the execution of the Contract – whether during developing phase or after their completion and whether before or after the termination, abandonment or breach of the Contract – the parties shall seek to resolve any such dispute or difference by mutual diligent negotiations in good faith within 14 (fourteen) days following a notice sent by one Party to the other Party in this regard.

ii. At future of negotiation the dispute shall be resolved through mediation and mediator shall be appointed with the mutual consent of the both parties.

iii. At the event of failure of mediation to resolve the dispute relating to this contract such dispute shall finally be resolved through binding Arbitration in accordance with GCC clause 38.

iv. The cost of the mediation and arbitration shall be shared by the parties in equal proportion however the both parties shall bear their own costs and lawyer's fees regarding their own participation in the mediation and arbitration. However, the Arbitrator may make an award of costs upon the conclusion of the arbitration making any party to the dispute liable to pay the costs of another party to the dispute.
v. Arbitration proceedings as mentioned in the above clause regarding resolution of disputes may be commenced prior to, during or after completion of the contract.

Notwithstanding any reference to the arbitration herein, the parties shall continue to perform their respective obligations under the Contract unless they otherwise agree that the Authority shall pay the Bidder any monies due to the Bidder.

Rules of procedure for arbitration proceedings:

Any dispute between the Authority and a Bidder who is a national of the Islamic Republic of Pakistan arising in connection with the present Contract shall be referred to adjudication or arbitration in accordance with the laws of the Islamic Republic of Pakistan including Arbitration Act 1940, however above provision shall prevail in referring the case to the Arbitrator.

Place of Arbitration and Award:

The arbitration shall be conducted in English language and place of arbitration shall be at Islamabad. The award of the arbitrator shall be final and shall be binding on the parties.

PUBLIC PROCUREMENT REGULATORY AUTHORITY (PPRA) EPADS

Bid Securing Declaration

Form 9: Bid Securing Declaration

Date: *[insert date (as day, month and year)]*

Bid No.:***P200268***

Alternative No.: **[insert identification No if this is a Bid for an alternative]**

To: **NUMS Procurement Directorate (National University Of Medical Sciences (NUMS)), Office Assistant ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province).**

We, the undersigned, declare that:

We understand that, according to your conditions, Bids must be supported by a Bid Securing Declaration.

We accept that we will be blacklisted and henceforth cross debarred for participating in respective category of public procurement proceedings for a period of (not more than) six months, if fail to abide with a bid securing declaration, however without indulging in corrupt and fraudulent practices, if we are in breach of our obligation(s) under the Bid conditions, because we:

1. have withdrawn or modified our Bid during the period of Bid Validity specified in the Form of Bid;

2. Disagreement to arithmetical correction made to the Bid price; or

3. having been notified of the acceptance of our Bid by the Procuring Agency during the period of Bid Validity, (i) failure to sign the contract if required by Procuring Agency to do so or (ii) fail or refuse to furnish the Performance Security or to comply with any other condition precedent to signing the contract specified in the Bidding Documents.

We understand this Bid Securing Declaration shall expire if we are not the successful

Bidder, upon the earlier of (i) our receipt of your notification to us of the name of the successful Bidder; or (ii) twenty-eight (28) days after the expiration of our Bid.

Signed: [insert signature of person whose name and capacity are shown] In the capacity of
[insert legal capacity of person signing the Bid Securing Declaration]

Name: [insert complete name of person signing the Bid Securing Declaration]

Duly authorized to sign the Bid for and on behalf of: [insert complete name of Bidder]

Dated on day of , [insert date of signing]
Corporate Seal (where appropriate)

Contract Form

SECTION IX: CONTRACT FORMS

THIS AGREEMENT made the _____ day of _____ 20_____ between **NUMS Procurement Directorate (National University Of Medical Sciences (NUMS)), Office Assistant ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province).**

(hereinafter called "the Procuring Agency") of the one part and *[name of Bidder]* of *[city and country of Bidder]* (hereinafter called "the Bidder") of the other part:

WHEREAS the Procuring Agency invited Bids for provision of goods, viz., **MS/NDMB&G/2026/42 Procurement of Electro-medical Lab equipment for Bio-genetics Lab-NIASR (NUMS) (*P200268*)** and has accepted a Bids by the Bidder for the provision of Goods in the sum of *[contract price in words and figures]* (hereinafter called "the Contract Price").

NOW THIS CONTRACT WITNESSETH AS FOLLOWS:

1. In this Contract words and expressions shall have the same meanings as are respectively assigned to them in the Conditions of Contract referred to.

2. The following documents shall be deemed to form and be read and construed as part of this Contract, In the event of any ambiguity or conflict between the Contract Documents listed below, the order of precedence shall be the order in which the Contract Documents are listed below:-

1. This form of Contract;
2. the Form of Bids and the Price Schedule submitted by the Bidder;
3. the Schedule of Requirements;
4. the Technical Specifications;
5. the Special Conditions of Contract;
6. the General Conditions of the Contract;
7. the Procuring Agency's Letter of Acceptance; and
8. [*add here: any other documents*]

3. In consideration of the payments to be made by the Procuring Agency to the Bidder as hereinafter mentioned, the Bidder hereby covenants with the Procuring Agency to provide the Goods related services and to remedy defects therein in conformity in all respects with the provisions of the Contract.

4. The Procuring Agency hereby covenants to pay the Bidder in consideration of the provision of Goods and the remedying of defects therein, the Contract Price or such other sum as may become payable under the provisions of the contract at the times and in the manner prescribed by the contract.

IN WITNESS whereof the parties hereto have caused this Contract to be executed in accordance with their respective laws the day and year first above written.

Signed, sealed, delivered by _____the _____ (for the Procuring Agency)

Witness to the signatures of the Procuring Agency:

……

Signed, sealed, delivered by _____the _____ (for the Procuring Agency)

Witness to the signatures of the Bidder: ……

PUBLIC PROCUREMENT REGULATORY AUTHORITY (PPRA) EPADS

Integrity Pact

Integrity Pact

DECLARATION OF FEES, COMMISSION AND BROKERAGE ETC. PAYABLE BYTHE SUPPLIERS OF GOODS, SERVICES & WORKS IN CONTRACTS WORTH RS.10.00 MILLION OR MORE

Contract Number: Contract Value: Contract Title:

Dated:

[Name of Supplier] hereby declares that it has not obtained or induced the procurement of any contract, right, interest, privilege or other obligation or benefit from Government of Pakistan or any administrative subdivision or agency thereof or any other entity owned or controlled by it (GoP) through any corrupt business practice.

Without limiting the generality of the foregoing [Name of Supplier] represents and warrants that it has fully declared the brokerage, commission, fee etc. paid or payable to anyone and not given or agreed to give and shall not give or agree to give to anyone within or outside Pakistan either directly or indirectly through any natural or juridical person, including its affiliate, agent, associate, broker, consultant, director, promoter, shareholder, sponsor or subsidiary, any commission, gratification, bribe, finder's fee or kickback, whether described as consultations fee or otherwise, with the object of obtaining or inducing the procurement of a contract, right, interest, privilege or other obligation or benefit in whatsoever form from GoP, except that which has been expressly declared pursuant hereto.

[Name of Supplier] certifies that it has made and will make full disclosure of all agreements and arrangements with all persons in respect of or related to the transaction with GoP and has not taken any action or will not take any action to circumvent the above declaration, representative or warranty.

[Name of Supplier] accepts full responsibility and strict liability for making and false declaration, not making full disclosure, misrepresenting fact or taking any action likely to defeat the purpose of this declaration, representation and warranty. It agrees that any contract, right interest, privilege or other obligation or benefit obtained or procured as aforesaid shall, without prejudice to any other right and remedies available to GoP under any law, contract or other instrument, be voidable at the option of GoP.

Notwithstanding any rights and remedies exercised by GoP in this regard, [Name of Supplier] agrees to indemnify GoP for any loss or damage incurred by it on account of its corrupt business practices and further pay compensation to GoP in an amount equivalent to ten time the sum of any commission, gratification, bribe, finder's fee or kickback given by [Name of Supplier] as aforesaid for the purpose of obtaining or inducing the procurement of any contract, right, interest, privilege or other obligation or benefit in whatsoever form from GoP.

Performance Guarantee Form

Performance Guarantee Form

To: **NUMS Procurement Directorate (National University Of Medical Sciences (NUMS)), Office Assistant ATR Building, Kutchery Chowk., Rawalpindi, Rawalpindi (District), Rawalpindi Division (Division), Punjab (Province).**

WHEREAS *[name of Bidder]* (hereinafter called "the Bidder") has undertaken, in pursuance of Contract No. *[reference number of the contract]* dated *[insert date]* for provision of Goods*(*hereinafter called "the Contract").

AND WHEREAS it has been stipulated by you in the said Contract that the Bidder shall furnish you with a Bank Guarantee by a reputable bank for the sum specified therein as security for compliance with the Bidder's performance obligations in accordance with the Contract.

AND WHEREAS we have agreed to give the Bidders guarantee:

THEREFORE, WE hereby affirm that we are Guarantors and responsible to you, on behalf of the Bidder, up to a total of *[amount of the guarantee in words and figures]*, and we undertake to pay you, upon your first written demand declaring the Bidder to be in default under the Contract and without cavil or argument, any sum or sums within the limits of *[amount of guarantee]* as aforesaid, without your needing to prove or to show grounds or reasons for your demand or the sum specified therein.

This guarantee is valid until the: *[insert date]*

Signature and seal of the Guarantors

_____

[name of bank or financial institution]

_____

[address]

_____

[date}

PUBLIC PROCUREMENT REGULATORY AUTHORITY (PPRA) EPADS

Annexure

No Annexure Defined.

PUBLIC PROCUREMENT REGULATORY AUTHORITY
(PPRA)
EPADS

Procurement Forms

PUBLIC PROCUREMENT REGULATORY AUTHORITY
(PPRA)
EPADS

PUBLIC PROCUREMENT REGULATORY AUTHORITY
(PPRA)
EPADS